


Luzchem Exposure Standard: LES-WPI-GR2-23

General information : Luzchem Research, Inc. produces and freely distributes these standards as a service to scientists involved in research in photochemistry, photobiology and photostability. These standards are available on Luzchem's website so that references to any of our standards can be used to define exposure conditions and should allow other scientists to replicate irradiation conditions readily. Luzchem has measured these exposure standards using a NIST traceable spectroradiometer. To our knowledge, spectral information is accurate within the experimental uncertainty of 1 nm.

Spectral Information: WPI-GR2, Phosphor Green, 550nm peak



Product Class	Well Plate Illuminator
Product Identifier	WPI-GR2
Spectral Range Monitored	220nm to 1100nm
Temperature	22.3 °C
Date of Measurement	March 30, 2023
Full Width Half Max	111.5 nm
Measured Peak Wavelength	559 nm
Irradiance Measured*	629 ± 34 Watts/m ²
Sample Power (12mm diameter spot)*	71.2 milliwatts

* measurement taken ~11.5mm above the surface of the WPI sample window, with a detector area 12mm in diameter centred in a well
 ± Accounts for minor variability between different Wells.

Phone 01-613-749-2442
 Fax 01-613-749-2393
 Toll Free 1-800-397-0977
 sales@luzchem.com

Luzchem Research, Inc.



WWW.LUZCHEM.COM

5509 Canotek Rd.
 Unit 12 & 13
 Ottawa, Ontario
 Canada, K1J 9J9