

**National Institute for Occupational Safety and Health
Respirator Branch
Test Data Sheet**



Task Number: TN-21384

Reference No.: CFR 84.180

Test: Inhalation Resistance Test

STP No.: 7

Manufacturer: Xiantao Zhongyi Safety Protection Products Co.,

Item Tested: ZYB-11

Filter Type: Filter Only

Sample	Maximum Allowable Resistance (MM of H2O)	Actual Resistance (MM of H2O)	Result
	Inhalation	Inhalation	
1	35	10.6	PASS
2	35	10.9	PASS
3	35	11.5	PASS

Overall Result: PASS

Signature:

Date: 3/20/2017

Engineering Technician

Task Number: TN-21384

Reference No.: CFR 84.180

Test: Inhalation Resistance Test

STP No.: 7

Manufacturer: Xiantao Zhongyi Safety Protection Products Co.,

Item Tested: ZYB-11

Comments:

Samples were tested on manometer 000285.

Was all equipment verified to be in calibration throughout all testing?	<input checked="" type="radio"/> Yes	<input type="radio"/> No
---	--------------------------------------	--------------------------

Signature:



Date: 3/20/2017

Engineering Technician

Task Number: TN-21384

Reference No.: CFR 84.181

Test: Sodium Chloride (NaCl) - N95

STP No.: 59

Manufacturer: Xiantao Zhongyi Safety Protection Products Co.,

Item Tested: ZYB-11

Comments:

Samples 1-10 were tested on TSI 8130 machine 000332 using timer 000213. Samples 11-15 were tested on TSI 8130 machine 000333. Samples 16-20 were tested on TSI 8130 machine 000334.

Was all equipment verified to be in calibration throughout all testing? Yes No

Signature:



Date: 3/21/2017

Engineering Technician