





Date Issued: 26 August 2008

CERTIFICATE No.: F690501/SP-EMG000200

## CERTIFICATE OF EMC COMPLIANCE

Order No.

G-44-2008-02096

Product submitted ....

Photon Compact Mat

Model: HM-2008PCM

Applicant (Manufacturer)....

HYUNDAI MEDICAL

213, Maesan-ri, Kwang-ju city, Kyungki-Do, Korea

Testing Laboratory .....

SGS Lab. 23 - SGS Testing Korea Co., Ltd.

Test Report Number(s)

F690501/RF-EMG002160

Date of Receipt

10 July 2008

Specification Requested ...:

EN 55014-1:2006, EN 61000-3-2: 2000/A2:2005

EN 61000-3-3: 1995/A1:2001/A2:2005, EN 55014-2:1997/A1:2001 (Category II)

## Conclusion

The apparatus meets the requirements of the above standards and hence fulfills the requirements of Directive 2004/108/EC.

This certificate is only valid for the equipment and configuration described, in conjunction with the test data detailed above. It does not permit the use of the SGS PRODUCT CERTIFICATION MARK.

The CE mark as shown below can be used, under the responsibility of the manufacturer, after completion of an EC Declaration of Conformity and compliance with all relevant EC Directives.

WonWoo Lee Manager

Eric Lee General Manager

Copyright of this certificate is owned by SGS Testing Korea and may not be reproduced other than in full and with the prior approval of the General Manager of SGS Testing Korea.

Member of SGS Group (Société Générale de Surveillance)

This document is issued, on the Client's behalf, by the Company under its General Conditions of Service printed overleaf.

SGSPAPER 06400800



F690501

The Client's attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.

her holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations