# **Clinical Trial Report**

A single-center, randomized, single-blind, retrospective, and confirmatory trial aimed to evaluate clinical performance of a study group (Ezplex® 2019-nCoV Real-time PCR Kit) which is devised for qualitative testing of SARS-CoV-2 virus using RNA extracted from oropharynx and nasopharynx specimens of patients suspected of coronavirus infection-19(COVID-19).

Study Protocol No.: GN-GCP-10

VERSION NO. 0.0

Document drafted on: March 18, 2020

Document completed on: March 18, 2020



Page 1 / 20

## **Revision History of Protocol Synopsys**

Revision No.	Date of Revision	Description of Revision	Drafted by
00	Mar. 18, 2020	Initial Release	KIM, Hyeong-nyeon

Trial Overview				
Protocol Title	A single-center, randomized, single-blind, retrospective, and confirmatory trial aimed to evaluate clinical performance of a study group (Ezplex® 2019-nCoV Real-time PCR Kit) which is devised for qualitative testing of SARS-CoV-2 virus using RNA extracted from oropharynx and nasopharynx swab specimens of patients suspected of coronavirus infection-19(COVID-19).			
Trial Design	Cross-over, randomized	d, single-blind		
Trial Objective	To assess the efficacy Ezplex® 2019-nCoV Re	of the investigational eal-time PCR Kit which	device by evaluating qualitatively detects	g clinical performance of the SARS-CoV-2 virus.
	Remaining specimens	of patients who were	referred to SM Lab	, tested for SARS-CoV-2
Subjects	Category	Registered no. of specimens	No. of excluded specimens	No. of specimens used for the trial
	Oropharynx swab	27	0	27
	Nasopharynx swab	26	0	26
	1			
	Manufacturer	Device Name/C	lassification No. (G	rade) / Model Name
Study Group	SML Genetree Co. Ltd.	IVD reagents for inf transmitted diseas other than 'high moderate infe Ezplex(	ectious disease mar e, Legally designate n risk pathogens', In ectivity), nucleic acid ® 2019-nCoV Real-t	ker(Diagnosis of Sexually d infectious pathogens fectious agents with test/D06080.01[3] ime PCR Kit
Control Group	Kogene Co. Ltd.	Kogene Co. Ltd.       IVD reagents for infectious disease marker(Diagnosis of Sexually transmitted disease, Legally designated infectious pathogens other than 'high risk pathogens', Infectious agents with moderate infectivity), nucleic acid test/D06080.01[3]         Powerchek™ 2019-nCoV Real-time PCR kit		
Sponsor	AHN, Ji-hoon, Representative of SML Genetree			
Address	SML Genetree Co. Ltd. (225, Baumoe-ro, Seocho-gu, Seoul)			
Trial Period	March 16, 2020 – march 18, 2020			
Participating Site	SM Lab (Molecular Bio	logy Testing Team)		
Address	Samgwang Bldg., 57, B	aumoe-ro 41-gil, Seoc	ho-gu, Seoul	

	Name	KIM, Hyeong-nyeon		
Principal Investigator	Department	SM Lab		
	Title	Specialized Medical Doctor		
Inclusion Criteria	1) Remaining speciment virus with an approve	s of patients who were referred to SM Lab, tested for SARS-CoV-2 d testing kit, and notified either positive or negative.		
	2) Remaining specimens	s kept frozen under -20 C and elapsed less than 2 months.		
	1) D			
	1) Remaining speciment	s that have been thawed and kept long in room temperature (19-		
Exclusion	25°C)			
Criteria	2) Remaining specimens less than 1,000 $\mu$ l, the minimum requirement.			
	3) Specimen containers are broken or contamination with other virus is suspected.			
	1. Primary Efficacy Assessment			
	The results of the re	emaining specimen are compared with the results from the		
Trial	investigational device to	evaluate the study group's clinical sensitivity.		
Methods	2. Secondary Efficacy As	ssessment		
	The results of the inve	stigational device are compared with the control group device to		
	evaluate the correlation	of the study group device and control group device.		
	26 nasopharynx specim	ens and 27 oropharynx specimens were collected according to the		
	inclusion criteria, and a	total of 53 specimens were tested with no exclusion.		
Irial Results	It was confirmed from	the test results that both the sensitivity and specificity were 100%,		
	and the Kappa value was 1.00 with the approved testing kit in terms of correlation.			

## March 18, 2020

Principal Investigator: KIM, Hyeong-nyeon (signature/seal)

## **Table of Contents**

1	Protocol Title	6
2	Name and Address of the Participating Site	6
3	Names and Titles of Principal Investigator, Investigators, and Co-investigators	6
4	Name and Title of Investigational Medical Device Manager	6
5	Name and Address of the Trial Sponsor	7
6	Inclusion/Exclusion Criteria for Trial Subjects and the Number of Target Subjects	7
19	9. Trial Period	8
8.	Trial Process of the Investigational Device	8
9.	Clinical Test Methods (Quantity of Use, Method of Use, Period of Use,	
	Combination Therapies, etc.)	15
10	) Clinical Trial Results	17
11	. Conclusion	20
12	2. Annex	20

## 1 Protocol Title

A single-center, randomized, single-blind, retrospective, and confirmatory trial aimed to evaluate clinical performance of a study group (Ezplex® 2019-nCoV Real-time PCR Kit) which is devised for qualitative testing of SARS-CoV-2 virus using RNA extracted from oropharynx and nasopharynx specimen of patients suspected of coronavirus infection-19(COVID-19).

## 2 Name and Address of the Participating Site

Category	Name	Address	Phone Number
Clinical Trial Institution	SM Lab	57, Baumoe-ro 41-gil, Seocho-gu, Seoul	+82-2-3497-5100

## 3 Names and Titles of Principal Investigator, Investigators, and Co-investigators

## 3.1 Principal Investigator and Investigators

	Name	Department	Title	Contact Number
Principal Investigator	KIM, Hyeong- nyeon	Testing HQ	Specialized Medical Doctor	+82-2-3497-5132
Investigator A	LEE, In-seob	Molecular Biology Testing Team	Team Head	+82-2-3497-5232
Investigator B	PARK, Joo-hyeon	Molecular Biology Testing Team	Team Staff	+82-2-3497-5232

## 3.2 Trial Statistics Personnel

Name	Department	Title	Contact Number
KIM, Hyeong-nyeon	, Hyeong-nyeon SM Lab (Testing HQ)		+82-2-3497-5132

## 4 Name and Title of Investigational Medical Device Manager

	Name	Department	Title	Contact Number
Study Personnel	LEE, In-seob	Molecular Biology Testing Team	Team Head	+82-2-3497-5232

## 5 Name and Address of the Trial Sponsor

#### 5.1 Trial Sponsor

Company	Name	Address	Phone Number
SML Genetree Co. Ltd.	AHN, Ji-hoon	225, Baumoe-ro, Seocho-gu, Seoul	+82-70-7425-3958

## 5.2 Monitoring Staff

Company	Name	Address	Phone Number
SML Genetree Co. Ltd.	LEE, Seung-Mok	225, Baumoe-ro, Seocho-gu, Seoul	+82-70-7425-3953

## 6 Inclusion/Exclusion Criteria for Trial Subjects and the Number of Target Subjects

#### 6.1 Inclusion/Exclusion criteria for trial subjects

#### 1) Inclusion criteria

- Remaining specimens of patients who were tested for SARS-CoV-2 virus with an approved testing kit\* in SM Lab and notified either positive or negative.
- Remaining specimens kept frozen under -20°C.

#### 2) Exclusion Criteria

- Remaining specimens that have been thawed and kept long in room temperature (19-25°C).
- Remaining specimens less than 1,000  $\mu l$ , the minimum requirement.
- Specimen containers are broken or contamination with other virus is suspected.

#### \* Powerchek<sup>™</sup> 2019-nCoV Real-time PCR kit (Kogene Biotech, KCDC Announcement No. 2020-105)

## 6.2 The number of collected subjects

Specimen Type	Positive	Negative
Oropharynx Swab	5	22
Nasopharynx Swab	5	21

## 19. Trial Period

Test Period	March, 2020			
lest renou	16th	17th	18th	
Clinical Observation and Trial				
Conduct				
Result Analysis and Statistical				
Processing				
Report Preparation				

## 8. Trial Process of the Investigational Device

## 8.1 Investigational Device Information

1. Device Name/Classification Number [Grade]: IVD reagents for infectious disease marker(Diagnosis of Sexually transmitted disease, Legally designated infectious pathogens other than 'high risk pathogens', Infectious agents with moderate infectivity), nucleic acid test / D06080.01[3]

- 2. Type Name (Model Name) : Ezplex® 2019-nCoV Virus Real-time PCR Kit
- 3. Manufacturer: SML Genetree
- 4. Storage Condition: 20 ℃
- 5. Measurement Principles

The device's measuring process mainly consisted of three steps:

① Specimen pretreatment

RNA is extracted from oropharynx and nasopharynx swab specimens

② PCR amplification

SARS-CoV-2 virus-specific primer and probe are used for PCR amplification and the amplified device is detected with the probe's fluorescence.

③ Result Analysis

The analysis program included in the device is used for analysis.

## 8.1 Preparation for the trial

- 1) The reagents should be stored at -20±2 °C and should avoid repeated thawing and freezing.
- 2) The reagents should be completed melted before use.
- 3) Since DNA degradation is likely to happen in case of positive control reagents, it is recommended to divide the reagents into small amounts for 1-2 tests and keep them frozen.
- 4) CFX96 Real-time PCR detection system (Import License 10-205) is a necessary machine for the testing. Therefore, it should be turned on before the test for booting and should be kept in an idle condition.

## 8.2 Trial Process

1) Refer to the below table and prepare master PCR mixes according to the number of samples to be tested (See Tables 1 and 2).

Component	Volume
RQ Mixture	10
nCoV P+P	5
Total	15

Table 1. When IC is separately included in the specimen extraction (Unit: uL)

Table 2. When IC is not separately included in the specimen extraction (Unit: uL)

Component	Volume
RQ Mixture	10
nCoV P+P	5
IC	0.1
Total	15

- 2) Divide 15uL of the master PCR mix into tubes and divide 5uL of ribonucleic acid (RNA) extracted from specimens and mix them well.
- 3) Both positive and negative control reagents will be tested together for the accuracy of the test.

4) Set the configurations of the machine as follows:

Step	Temperature / Time	Cycle
	25 °C / 2 min	
Hold	50 °C / 30 min	1 Cycle
	95 ℃ / 5 min	
Guele	95 °C / 15 sec	
Cycle	60 °C / 45 sec	40 Cycles

#### 8.3 Result Analysis

#### A. Amplification setup

After the test is finished, use the machine software (CFX manager) to set up the threshold values of fluorescence for each detection target as follows: '1000' for FAM, 500 for HEX, '150' for Cy5 and '100' for Cy5.5 (Quasar705).

#### **B. Software installation**

(1) Files to be pre-installed

Before installing the analysis software, double-click the 'vc\_redist.x86.exe' file as shown in the below image to install the 'Microsoft Visual C++ 2015 Redistributable(x86)' in the same folder.

₿ Microsoft Visual C++ 2015 Redistributable (x86) - 14.0 —		×
Microsoft Visual C++ 2015 Redistributable (x86) - 14.0.23	026	
MICROSOFT 소프트웨어 사용권 계약서		^
MICROSOFT VISUAL STUDIO 2015 추가 기능, VISUAL STUDIO : 및 C++ 재배포 가능 패키지	SHELL	
본 사용권 계약은 Microsoft Corporation(또는 거주 지역에 따라 계일 	┋사 중 	~
♥설치(!)	닫기(C)	)

(2) After installing the MS Visual C++ 2015, click 'Run as Administrator' for 'Setup.exe' file in the Setup folder.



(3) After running the setup file, click the 'Install' button as shown in the below image and start installation.



(4) After the installation is complete, the execution file can be seen from the Windows Start menu.

Note: If the program is not executed, install the 'Microsoft Visual C++ 2015 Redistributable X64 version and reinstall the analysis program (vc\_redist.x64.exe).



#### C. Software-enabled analysis

(1) Check that PCR test is complete and transfer the test data to an excel spreadsheet through 'Export All Data Sheets' in the 'Tools' menu of CFX96 Manager software (Create a folder and save the file in it).



2) Run the analysis software (Genetree Viewer), click 'Open' on the upper left to locate the folder where the transferred excel spreadsheet is saved and open the file with a file name ending in 'Quantitation Amplification Results.'

4	이름	수정한 날짜	05 65 66	37
	🔒 결과	2019-04-04 오후	파일 볼더	
5	admin_2019-04-04 10-17-16_BR100841 - Allelic Discrimination Results	2019-04-04 오후	Microsoft Excel	11
	admin_2019-04-04 10-17-16_BR100841 - End Point Results	2019-04-04 오후	Microsoft Excel	34
	admin_2019-04-04 10-17-16_BR100841 - Gene Expression Results	2019-04-04 오후	Microsoft Excel	4
	admin_2019-04-04 10-17-16_BR100841 - Melt Curve Amplification Results	2019-04-04 오후	Microsoft Excel	666
	admin_2019-04-04 10-17-16_BR100841 - Melt Curve Derivative Results	2019-04-04 오후	Microsoft Excel	679
	admin_2019-04-04 10-17-16_BR100841 - Melt Curve Peak Results	2019-04-04 오후	Microsoft Excel	17
	admin_2019-04-04 10-17-16_BR100841 - Melt Curve Plate View Results	2019-04-04 오후	Microsoft Excel	16
	admin_2019-04-04 10-17-16_BR100841 - Melt Curve Summary	2019-04-04 오후	Microsoft Excel	9
	admin_2019-04-04 10-17-16_BR100841 - Quantitation Amplification Results	2019-04-04 오후	Microsoft Excel	280
	admin_2019-04-04 10-17-16_BR100841 - Quantitation Ct Results	2019-04-04 오후	Microsoft Excel	34
	admin_2019-04-04 10-17-16_BR100841 - Quantitation Plate View Results	2019-04-04 오후	Microsoft Excel	17
	Dadmin_2019-04-04 10-17-16_BR100841 - Quantitation Summary	2019-04-04 오후	Microsoft Excel	21
	¢			>
	01 ≣ (N):	v e	oport file (*.xlsx, *.xls, *	.csv) v
		Г	g7l(0)	취소

3) Click 'Please select a kit' on the upper part of the screen and select an appropriate kit (2019-nCoV real-time PCR:CFX96) for the experimented panel as shown in the below image.

		Open	Plea	se sele	ct a kit								→= Threshold
			H.Py	lori rea	l-time F	CR : CF	X96						
	1	2	MP-	dR real-	Ct								
A	?	?	2019	9-nCoV	real-tim	e PCR :	CFX96	;					17
в	?	?	2019	9-nCoV	(N) rea	l-time P	CR : CF	X96					0.9-
С	?	?	?	?	?	?	?	?	?	?	?	?	0.7 -
D	?	?	?	?	?	?	?	?	?	?	?	?	0.8 -
E	?	?	?	?	?	?	?	?	?	?	?	?	0.4 -
F	?	?	?	?	?	?	?	?	?	?	?	?	0.3 -
G	?	?	?	?	?	?	?	?	?	?	?	?	0.1 -
н	?	?	?	?	?	?	?	?	?	?	?	?	0

4) As shown in the below image, test results for each well are displayed according to the selected kit.

Note : See the below table for detailed of	description of the result screen .
--	------------------------------------

No.	Description
1	Positive/negative results for each well is indicated as '+' or '-,' respectively.
2	Ct and fluorescence values for each well's results are shown on a graph.
3	Ct values for each well's results are shown in numbers and qualitative results are displayed.
4	Analysis results are exported as an excel spreadsheet.

GEN	ETR	EE/\	NEWE	R 1.13	2.570															- B.
E		-	281	9400	/ real-Siz	N POR	CFX96	881279	antala Anglia	-	-	ē n	ushold					··· D	ExportToExes	6
	3	1	3	4	5		2 8		" "1	B(	Ct								04	
	+	1.									20	19							Q.	
	٠	Ŀ	•					1.0	• •		30	-								
Ð	٠					+	÷. •		+ 1		25	-						/		1.00
ŀ I	٠								4 1			1						1 -		VIC
	1	1	1	1		1	1 1		1 1		10	-					1	1		CV5
	1					1	1.1		1 1			-					1			
1	÷								1 1											
1		-	-	-	-	-	-	-							-					
	engle	0	Dary	nie –	we	Rea	Cut.	Conternel	ravi(tidtp)	VICIEI	CYSIN	Q705(C	DB NK	1. 198	Valuety	Comment				
	ົ	21			A01	A.	- 01	Unkin-D1	25.71	27.92	30.91	NA	Postive	Positive	valid .					
Ľ	ગ	-			801		01	Unke-80	2 25.57	27.58	30.95	NA	Positive	Postive	valid					
					C21	c	01	Univ-D	30.69	32.77	26.97	84/4	Positive	Postive	Valid					
F					001	D	01	Unke-Di	10.71	32.64	26.83	NA	Postive	Postive	blev					_
F					601	6	01	Unker-85	s NoA	N/6	NA	14.4	invalid	invalid	invalid					
F					F01	F	01	Unkn-De	N/A	NA	N/A	NU.	maid	insid	invalid					
					001	0	01	Unici-20	25.63	17.72	30.86	N/A	Postlys	Poster	Valid					
F					HET	н	01	Unico-Di	NOA I	N/06	N/A	14/4	invalid	invalid	invalid					
F					A02		02	Unico-Di	25.51	27.53	30.71	N/A	Postive	Poster	Valid					
F					802	8	02	Unio-10	25.28	27.22	30.47	NA	Postive	Postive	Valid					
	_																			
Sar	np	le I	USa	mp	le ¥	rell	How	COL	Conten	FAN	исна	Hp. VI	L(E)	C Y5(N)	Q 705(	U) 환경	MIXI	<b>완</b> 성	Validity	Comme
					A	.01	A	01	Unkn-01	N/A	1	N/	A	N/A	29,47	Neg	ative	Negative	Valid	
		-	-		B	01	В	01	Unkn-02	N/A	(s	N/	A	N/A	28,88	Neg	ative	Negative	Valid	
	1		1		E	12	E	12	Unkn-93	N/A	() ()	N/	A	N/A	28,66	Neg	ative	Negative	Valid	
		6			F	12	F	12	Unkn-94	14.7	2	16	40	18,16	N/A	Posi	tive	Positive	Valid	
Γ			T		G	i12	G	12	Unkn-95	31,6	6	33	30	38,31	N/A	Posi	tive	Positive	Valid	
					H	12	Н	12	Unkn-96	N/A	\	N/	Ά	N/A	N/A	Inva	id	Invalid	Invalid	

Page 13 / 20

FAM (RdRp) HEX (E)		Cy5 (N)	Quasar705 (IC)	Judgment*	Remarks
<40	<40	<40	Any	Positive	
<40	≥40 or Neg	<40	Any	Inconclusive*	
<40	<40	≥40 or Neg	Any	Inconclusive*	
≥40 or Neg	<40	<40	Any	Inconclusive*	
<40	≥40 or Neg	≥40 or Neg	Any	Negative**	
≥40 or Neg	<40	≥40 or Neg	Any	Negative**	
≥40 or Neg	≥40 or Neg	<40	Any	Negative**	
≥40 or Neg	≥40 or Neg	≥40 or Neg	<38	Negative	
≥40 or Neg	≥40 or Neg	≥40 or Neg	≥38 or Neg	Invalid	Retest after re- extraction

#### D. How to interpret the results

Results are determined by checking the Ct values according to the table below.

\* If positive results are confirmed for 2 or less genes from using this device, an additional test should be done by retesting with the same device or using another diagnostic method.

\*\* If only one type of gene is detected regardless of genetic types, the result must be judged as negative.

## 8.4 Severity

- Positive control should be included in the test, and the test result should be positive for all the test targets.
- D.W. which is used as negative control should be included in the test, and the test result should be negative in all kits. If any positive result is shown in the negative control well, a carry-over or contamination can be suspected and a retest should be performed.

## 9. Clinical Test Methods (Quantity of Use, Method of Use, Period of Use,

## **Combination Therapies, etc.)**

## 1. Specimen Handling

- Collected and tested specimen were remaining oropharynx and nasopharynx swab specimens of anonymized (coded with unique identifiers) patients referred to the participating site (SM Lab) that were confirmed either positive or negative of SARS-CoV-2 virus using the existing test kits (See Table 3).
- Approved medical devices (See Table 3) were used to extract ribonucleic acid (RNA) from all the remaining specimens.
- The principal investigator took the responsibility to have all the specimens coded with unique identifiers and randomized so that an investigator may not become aware of any personal information. The principal investigator delivered the uniquely identified and randomized specimens to the investigator B and instructed to perform the tests.

## 2. Clinical Test

- The investigator B received and tested the specimens using the medical devices allocated for the clinical trial (See Table 3) according to the dosage and administration. Then, the results were delivered to the investigator A for interpretation.
- The investigator A used an exclusive software (Genetree Viewer 1.13.2.570) for the investigational device to interpret the test results, recorded the results on the CRF and informed the principal investigator and no retest was required.

Model Name	License Info	Manufacturer	Remarks
Powerchek™ 2019-nCoV Real-time PCR kit	Emergency Use Authorized device as per KCDC Announcement No. 2020-105	Kogene Biotech	Control group
CFX96 Real-time PCR	Import License 10 - 205	Bio-Rad	Study group equipment
Nextreactor NX-48	Gyeong-in Jesin14-66	Genolution	Extraction of ribonucleic acid

## [Table 3]. Medical devices that were used together with the investigational device

#### 3. Result Analysis

The principal investigator received the study group results from the investigator A and compared the results with the existing specimen results to assess clinical sensitivity and specificity according to Table 4. No sequencing was required due to discrepancy.

Remaining	Investigational	Retest	Sequencing results	Final
specimen results	device results	results	Sequencing results	Judgment
	Positive	-	-	Positive
	Negative	-	Positive	False Negative
	5		Negative	Negative
Positive		Positive	-	Positive
	Invalid	Negative	Positive	False Negative
		5	Negative	Negative
		Invalid	-	Invalid
	Positive	-	-	False Positive
	Negative	-	-	Negative
Negative		Positive	-	False Positive
	Invalid	Negative	-	Negative
		Invalid	-	Invalid

#### [Table 4]. Methods for result interpretation

## **10 Clinical Trial Results**

## 10.1 Summary of the clinical trial targets

1) A summary of the clinical trial targets can be in the table below.

Specimen Selection	Remaining specimens of patients who were referred to SM Lab, tested for SARS-	
specifien selection	CoV-2 virus with an existing test kit, and notified either positive or negative.	
Number of	F2 mm	
Screened Specimen	53 Cases	
Number of	F2 mm	
Analytes	53 Cases	
Excluded Specimen	None	
Final Number of		
Tested Specimen	53 cases	

## 10.2 Statistical Analysis

#### 10.2.1 Efficacy Assessment

1) Regarding efficacy parameters available for the assessment, ratios are presented for specimens that were positive before but judged negative with the investigational device or specimens that were negative before but judged positive with the investigational device

2) Original specimen results are compared with the results from the investigational device and presented in a table.

#### 10.2.2 Safety Assessment

Mechanical issues or malfunction/defect and their percentages are suggested for the investigational device as well as other medical devices that were used together.

#### 10.2.3 Analysis Sets

Analysis sets for this trial are defined as remaining oropharynx and nasopharynx swab specimens that satisfy the 'inclusion and exclusion criteria.'

#### 10.2.4 Statistical Analysis Method

MedCalc Statistical Software version 14.8.1 (MedCalc Software bvba, Ostend, Belgium; http://www.medcalc.org; 2014) was used for statistical analysis.

In order to assess clinical performance of the investigational device in comparison with the specimens' original results, clinical sensitivity and specificity were calculated. The Kappa value was also calculated to evaluate the correlation with a kit which was approved earlier.

## 10.3 Trial Results

#### 10.3.1 The number of specimens that were judged positive or negative after testing.

Number of specimens for completed testing							
Specimen Type Positive Negative							
Oropharynx swab	5	22					
Nasopharynx swab	5	21					

#### 10.3.2 The number of specimens with discrepant results (positive/negative)

Category	Original result	Study Group Result	No. of discrepancy	Concordance (%)
Positive	10	10	0	100
Negative	43	43	0	100

# 10.3.3 Mechanical issues or malfunction/defects for medical devices used together with the investigational device.

There was no issue regarding degeneration of diagnosis reagents or malfunction/defects from the devices used in the course of conducting this trial.

## 10.3.4. Compliance with the inclusion/exclusion criteria

Category	No. of Specimen	Compliance with the inclusion/exclusion criteria		
Positive	10	100 %		
Negative	43	100 %		

#### 10.3.5. Test Results

#### 1. Primary Efficacy Assessment Results

(1) Overall results

SARS-CoV-2 Virus		Control Group (Specimens and Sequencing Results)				
		Positive	Negative			
Study Positive		10	0			
Group	Negative	0	43			
Sensitivity (95% CI)		100 % (69.2 ~ 100 %)				
Specificity (95% CI)		100 % (91.8 ~ 100 %)				

## (2). Results by Specimen Type

Oropharynx Swab Positive Study Group Negative	Control Group			Nacanhan <i>u</i> ny		Control Group					
Oropharynx Swab		Positive	Negative	Total	Nasopharynx		Positive	Negative	Total		
	Positive	5	0	5	Study Group			Positive	5	0	5
Study Group	Negative	0	22	22		Negative	0	21	5 21 26		
•	Total	5	22	27		Total	5	21	26		
Sensitivity (95% CI)		100 % (47.8 ~ 100 %)			Sensitivit	y (95% CI)	100 9	% (47.8 ~ 10	0 %)		
Specificit	y (95% CI)	100 % (84.6 ~ 100 %)		Specificity (95% CI)		100 % (83.9 ~ 100 %)					

## 2. Secondary Efficacy Assessment Results

## (1) Overall Results

SARS-CoV-2 Virus		Control Group (Specimens and Sequencing Results)				
		Positive	Negative			
Study		10	0			
Group	Negative	0	43			
Positive Percent Agreement		100 % (69.2 ~ 100 %)				
Negative Percent Agreement		100 % (91.8 ~ 100 %)				
Kappa value		1.00				

Oropharynx Swab		Control Group			Nacanhan my Such		Control Group		
		Positive	Negative	Total	wasopharynx Swab		Positive	Negative	Total
Positive	Positive	5	0	5	<b>C</b> ( <b>1</b>	Positive	5	0	5
Study	Negative	0	22	22	Study	Negative	0	21	21
Group	Total	5	22	27	Group	Total	5	21	26
Positive Percent Agreement		100 % (47.8 ~ 100 %)		Positive Percent Agreement		100 % (47.8 ~ 100 %)			
Negative Agree	e Percent ement	100	100 % (84.6 ~ 100 %)		Negative Percent Agreement		100 % (83.9 ~ 100 %)		
Карра	a value		1.00		Kappa value		1.00		

(2) Results by Specimen Type

## 11. Conclusion

Clinical sensitivity and specificity of Ezplex® 2019-nCoV Real-time PCR Kit was evaluated in order to assess the device's clinical performance by testing remaining specimens of those that were tested with an already approved test kit and comparing the results. This device is for an in-vitro diagnostic test which uses oropharynx and nasopharynx swabs collected from people suspected of coronavirus infection to extract ribonucleic acid. Therefore, the device is judged to have no safety issues since it does not cause a direct harm on human body. 53 specimens that were involved in the trial all met the inclusion criteria and none were excluded.

According to the result of testing the study group device for the targeted SARS-CoV-2 Virus, a high clinical performance was confirmed from the study group. The high clinical performance of the study group was validated from 100% sensitivity and specificity of the study group that were determined from the collected specimens, 100% positive/negative percent agreement in terms of correlation, and the Kappa value of 100.

## 12. Annex

- Case Report Form

2	퇴기관	(관 ㈜에스옘엛제니트리			기관	삼광의료재단	
No.	Sample Code	겹체	대초군 결과	FAM (1.000)	HEX (500)	Cy5 (150)	Quasar75 (100)
1	SML-CV2-01	구인두	Positive	31.32	24.58	25.98	35.18
2	SML-CV2-02	구연두	Negative	N/A	N/A	N/A	33.51
3	SML-CV2-03	비연두	Positive	34.90	28.74	29.88	36.19
4	SML-CV2-04	비언두	Negative	N/A	N/A	N/A	32.99
5	SML-CV2-05	구연두	Negative	N/A	N/A	N/A	32.58
6	SML-CV2-06	구인두	Negative	N/A	N/A	N/A	32.30
7	SML-CV2-07	비인두	Negative	N/A	N/A	N/A	32,44
8	SML-CV2-08	비안두	Positive	25.27	23.16	22.19	34.36
9	SML-CV2-09	비안두	Negative	N/A	N/A	N/A	32.57
10	SML-CV2-10	비안두	Negative	N/A	N/A	N/A	32.70
11	SML-CV2-11	비안두	Negative	N/A	N/A	N/A	32.30
12	SML-CV2-12	비안두	Negative	N/A	N/A	N/A	32.64
13	SML-CV2-13	비연두	Negative	N/A	N/A	N/A	32.01
14	SML-CV2-14	구이동	Negative	N/A	N/A	N/A	32.07
15	SML-CV2-15	비외동	Negative	N/A	N/A	N/A	32.54
16	SML-CV2-16	그이드	Negative	N/A	N/A	N/A	32.63
17	SML-CV2-17	비아트	Positive	15.85	27.99	28.33	34.26
19	SML-CV2-17	그이트	Negative	53.05	67.99 N/A	20.35 N/A	24.27
10	SML-CV2-10	2015	Desitive	21.05	25.25	29.57	22.50
20	SML-CV2-19	205	Mogative	51.55	AL/A	20.37	22.44
20	SIVIL-CV2-20	2015	Negative	N/A	DI//S	N/A	22.64
21	SIML-CV2-21	비아드	Negative	NVA	N/A	N/A	32.04
22	SML-LV2-22	비인주	Negative	N/A	N/A	IN/A	32.23
23	SML-LV2-23	미인주	Negative	N/A	N/A	N/A	33.82
24	SML-CV2-24	727	Negative	N/A	N/A	IN/A	33.19
25	SML-CV2-25	구인두	Negative	N/A	N/A	N/A	32.71
26	SML-CV2-26	미인두	Negative	N/A	N/A	N/A	32.20
27	SML-CV2-27	비인수	Negative	N/A	N/A	N/A	32.37
28	SML-CV2-28	비인두	Negative	N/A	N/A	N/A	32.22
29	SML-CV2-29	비인두	Negative	N/A	N/A	N/A	32.54
30	SML-CV2-30	구인두	Negative	N/A	N/A	N/A	32.42
31	SML-CV2-31	구인두	Negative	N/A	N/A	N/A	33.70
32	SML-CV2-32	비인두	Negative	N/A	N/A	N/A	32.01
33	SML-CV2-33	구인두	Negative	N/A	N/A	N/A	32.10
34	SML-CV2-34	구인두	Negative	N/A	N/A	N/A	31.20
35	SML-CV2-35	비인두	Negative	N/A	N/A	N/A	32.39
36	SML-CV2-36	비인두	Positive	35.25	30.77	30.99	32.34
37	SML-CV2-37	구인두	Negative	N/A	N/A	N/A	32.11
38	SML-CV2-38	구인두	Negative	N/A	N/A	N/A	32.72
39	SML-CV2-39	구인두	Positive	30.07	26.55	26.28	35.08
40	SML-CV2-40	비인두	Negative	N/A	N/A	N/A	32.78
41	SML-CV2-41	구인두	Negative	N/A	N/A	N/A	33.84
42	SML-CV2-42	비인두	Negative	N/A	N/A	N/A	32.62
43	SML-CV2-43	구인두	Negative	N/A	N/A	N/A	32.03
44	SML-CV2-44	구인두	Negative	N/A	N/A	N/A	32.50
45	SML-CV2-45	구인두	Positive	32.25	27.59	27,59	32.60
46	SML-CV2-46	구인두	Negative	N/A	N/A	N/A	32.05
47	SML-CV2-47	비안두	Negative	N/A	N/A	N/A	32.01
48	SML-CV2-48	비인두	Positive	27.09	24.57	23.40	32.64
49	SML-CV2-49	비인두	Negative	N/A	N/A	N/A	32.46
50	SML-CV2-50	구인두	Positive	28.11	27.03	25.74	32.91
51	SML-CV2-51	구인두	Negative	N/A	N/A	N/A	32.13
52	SML-CV2-52	구인두	Negative	N/A	N/A	N/A	32.60
53	SML-CV2-53	비인두	Negative	N/A	N/A	N/A	32.36
	중례기록서의	모든 내용을	검토하였으며, #	바짐없이 정확히	기록하였음을	· 확인핰니다	