



Seal&Cut MV1

Autoclaving Instructions



Warning! Ignoring the following instructions may lead to damaging the tools or shortening their life span.

- ➔ **Please make sure that the vessel sealing instrument is open (not closed) during autoclaving**
- ➔ **Recommended autoclave cycle for all accessories:**
 - **Sterilization** temperature: 250°F, maximum time – 20 minutes
 - **Drying** temperature: 195°F, maximum time – 10 minutes