



STATE OF NEW YORK
DEPARTMENT OF AGRICULTURE AND MARKETS
10B Airline Drive, Albany, New York 12235
1-800-554-4501
www.agmkt.state.ny.us

Certificate of Approval
for Weighing and Measuring Devices

New York State Certificate Number: 8766
Effective Date: December 3, 2008

NTEP Certificate of Conformance Number: 97-087A2

For:


Bench & Counter Scale
Digital Electronic
Model: CCI-220 Series
 n_{max} : 3750
 e_{min} : 0.02 lb (0.01 kg)
Capacity: 75 – 300 lb (35 – 100 kg)
Platform: See below
Accuracy Class: III

Submitted By:

CCI Scale Company, Inc.
4350 Transport Street, #108
Ventura, CA 93003
Tel: (805) 658-0422
(800) 900-0224
Fax: (805) 658-0436
Contact: Tom Spring
Email: CCIwest@hotmail.com

This certifies that the items specified in the above National Type Evaluation Program (NTEP) Certificate of Conformance are hereby approved for sale or use in the State of New York.

The NTEP Certificate of Conformance, as issued by the National Conference on Weights and Measures, is accepted under the terms of 1NYCRR Part 220.1. Evaluation results and device characteristics necessary for inspection and use in commerce are stated in the NTEP Certificate of Conformance. Copies of the NTEP Certificate of Conformance are available on request and are available for inspection at the Bureau's Metrology Office at Room 123, Building 7A, State Office Campus, 1220 Washington Avenue, Albany, NY 12235.



Ross J. Andersen, Director
NYS Bureau of Weights and Measures