

## Introduction

- Dental hygiene ranks first of all U.S. occupations for prevalence of carpal tunnel syndrome, musculoskeletal diseases (MSDs) & upper extremity disorders.<sup>1</sup>
- In one study, 64-96% of hygienists experienced symptoms of MSDs within a 12-month period.<sup>2</sup>
- 2/3 of dental clinicians report occupational musculoskeletal pain.<sup>3</sup>
- One third of dental clinicians retire early due to MSDs.<sup>4</sup>
- Ultrasonic scaling, and many forms of dental instrumentation are related to a wide range of musculoskeletal diseases, as well as intra- and postoperative discomfort and fatigue.<sup>5</sup>

**Goal: to evaluate the effect of a novel wearable cord-holding device on muscle work, fatigue, musculoskeletal symptoms and comfort related to ultrasonic scaling.**

## Materials and Methods

- Protocol granted exempt status by University of California Irvine IRB.
- Randomized, controlled, crossover study design.
- 5 hygienists served as testers: age 32-54 years; mean 41 years.
- 2 testers had 5-10 years of clinical experience; 3 testers 11-20+ years.
- Testers performed standardized ultrasonic scaling task twice: with and without use of wearable cord-holding device (Cordeze<sup>®</sup>, Veil Products, Phoenix, AZ 85087) (Figure 1). Cord-holder attaches to ultrasonic scaler cord to serve as stress-breaker for pullback.
- Using dental typodont with standardized calculus load, testers scaled each lingual or buccal surface of each quadrant for 2 minutes.
- Four wireless surface electromyography (sEMG) electrodes were attached to hands and arms of tester's dominant limb to measure activity in 4 muscles: *extensor digitorum communis*, *flexor digitorum superficialis*, *extensor carpi radialis brevis*, *first dorsal interosseous*.
- Evaluation criteria:
  1. Hand, wrist, arm fatigue & comfort: (visual analog scale (VAS) recorded immediately post-scaling; 0-10 scale; 0 best, 10 worst).
  2. Muscle work: sEMG traces analyzed using BTS EMG analyzer<sup>®</sup> software (FREEEMG, ©BTS Engineering, Quincy, MA).
  3. Cord pullback force: tensional dynamometer each site (N force).
  4. Efficacy: percent of each buccal or lingual quadrant surface scaled within 2 minutes

Statistical Analysis: sEMG trace data were analyzed using multivariate ANOVA and Bonferroni post-hoc tests; t- tests were used for the remaining analyses. Significance level set at p<0.05.

## Results and Discussion



Figure 1: sEMG electrodes and wristband cord-holder (white arrow) *in situ*

1. Combined mean fatigue in all 4 muscles was reduced by 60% using the wristband; mean comfort was improved by a factor of 3 (sig., p<0.05) (Figure 2).





Figure 2: Mean VAS Scores: 1= overall fatigue in hands, fingers, wrists, 2= comfort in thumb, 3= comfort in wrist, 4= comfort in fingers, 5= comfort in palms

2. Mean VAS score for each of the 5 evaluation categories was sig. better when cord-holder was used (p<0.05), demonstrating that user fatigue and comfort at all sites (thumb, wrist, finger, palms) were significantly improved during wristband cordholder use (Fig. 2).
3. Based on sEMG measurements, work/s during scaling was reduced by 30% and total work to complete the scaling task by 25% using the wristband (sig., p<0.05) (Fig. 3).



4. Hygienists registered significantly fewer complaints related to discomfort or pain (p<0.05) when working with the wristband (Table 1).
5. Testers also remained symptom-free during scaling for a significantly longer period of time (p<0.05) using the wristband (Table 2).
6. Testers recorded significantly fewer complaints (p<0.05) at each anatomical site when working with the wristband vs. without (Table 3).

## Results and Discussion (Continued)

	w/ Cordeze	w/o Cordeze
TG	4	12
LB	3	19
RS	0	14
RV	4	13
YH	4	14
<b>Total</b>	<b>15</b>	<b>72</b>
<b>Mean</b>	<b>3</b>	<b>14.4</b>

	w/ Cordeze	w/o Cordeze
TG	4.1	4.1
LB	3.05	0.05
RS	none	2.15
RV	8.07	2.37
YH	4.15	1.5
<b>Total</b>	<b>19.37</b>	<b>8.67</b>
<b>Mean</b>	<b>3.87</b>	<b>2.03</b>

	w/ Cordeze	w/o Cordeze
Purlicue	5	8
Wrist	2	4
Index finger	1	14
Palm	2	8
Forearm	1	14
Thumb	3	18
Hand	0	2
Upper arm and neck	1	4
<b>Total</b>	<b>15</b>	<b>72</b>

Table 1: Number of Complaints Table 2: Time until First Complaint Table 3: Total Number Complaints/Site

7. Cord pullback force was eliminated when scaler cord was attached to wristband, while measuring 2.3 N when wristband was not used (Fig 4).




Figure 4: Mean pullback force without wristband for 8 intraoral sites scaled: 1-URQ Buccal, 2-ULQ Buccal, 3-ULQ Lingual, 4-URQ Lingual, 5-LRQ Buccal, 6-LLQ Buccal, 7-LLQ Lingual, 8-LRQ Lingual. Pullback force at all 8 sites with wristband in place measured 0 N.

8. While wearing the wristband, all testers completed cleaning all surfaces during the given time allotment. Without the wristband, 1 hygienist did not complete scaling in 2 areas, and another hygienist failed to scale 1 surface within the 2-minute time allotment.

## Conclusions and Clinical Relevance

**The results of this pilot study indicate that a novel wristband cord-holder may improve ergonomics & reduce musculoskeletal burden of ultrasonic scaling while supporting efficient instrumentation.**

## Acknowledgements

We gratefully acknowledge funding from LAMMP NIH/NIBIB P41EB05890, Cordeze Inc., and the Arnold and Mabel Beckman Foundation.

## References

1. Leigh JP, Miller TR. Occupational illnesses within two national data sets. *Int J Occup Environ Health* 1998; 4(2):99-113
2. Marshall ED, Duncombe LM, Robinson RQ, et al. Musculoskeletal symptoms in New South Wales dentists. *Aust Dent J* 1997; 42(4):240-6
3. Auguston TE, Morken T. Musculoskeletal problems among dental health personnel. A survey of the public dental health services in Hordaland. *Tidsskr Nor Laegeforen* 1996; 116(23):2776-80
4. Akesson I, Johnsson B, Rylander L, et al. Musculoskeletal disorders among female dental personnel: clinical examination and a 5-year follow-up study of symptoms. *Int Arch Occup Environ Health* 1999; 72(6):395-403
5. Osborn JB, Newal KJ, Rudney JD, et al. Musculoskeletal pain among Minnesota dental hygienists. *J Dent Hyg* 1990; 64(3):132-138