



# MONARCH INSTRUMENT

## EU Declaration of Conformity

**We**

Monarch Instrument  
15 Columbia Drive  
03031, Amherst, NH  
(603) 883-3390  
[sales@monarchinstrument.com](mailto:sales@monarchinstrument.com)

**declare under our sole responsibility that the following product**

Equipment: Nova Pro 300, 500  
Brand name: Monarch Instrument  
Model/type: Measurement, Control Equipment and Laboratory Use

**is in conformity with the**

Low Voltage Directive (LVD) 2014/35/EU  
Electromagnetic Compatibility Directive (EMC) 2014/30/EU  
Restriction of Hazardous Substances (RoHS) Directive 2011/65/EU

**and the following harmonized standards and technical specifications have been applied:**

LVD: EN61010-1:2010 Safety Regulation

EMC: EN61326-1:2013 Class A  
IEC61000-4-2: 2008  
IEC61000-4-3: 2007

RoHS: EN 50581:2012

Laser IEC 60825-1:2014-3

### **Additional Information**

EMC Testing of this product was performed by Retilf Testion Laboratories, NH (File R-6104N).

June 7, 2016

Manufacturer: Monarch Instrument

---

Alan Woolfson, CTO Engineering (Authorized Signature)