



## **DECLARATION OF CONFORMITY**

As manufacturer, Monarch Instrument hereby declares under our sole responsibility that the product:


**Name: Remote Optical Laser Sensor**  
**Model: ROLS**

to which this declaration relates follows all the applicable essential requirements and other provisions of the European Council Directives listed when operated in accordance with the user guide. This product will bear the CE Mark label (**CE**).

### **DIRECTIVES:**

- **Council Directive 2012/19/EU** (WEEE Directive)
- **Council Directive 2015/863/EU** (RoHS 3 Directive)
- **Council Directive 2014/30/EU** (EMC Directive)
  - EN 61326-1:2013
- **Council Directive 2014/35/EU** (Low Voltage Directive)
  - Electrical Safety: EN 61010-1:2010
  - Laser Safety: EN 60825-1:2014

EMC testing of this product was performed by Retlif Testing Laboratories (File R-5568N-8).

  
\_\_\_\_\_  
Alan Woolfson  
Chief Technical Officer (Authorized Signature)

22 July 2019  
Date

Amherst, NH, USA  
Place

**Manufacturer's Address:**  
Monarch Instrument  
15 Columbia Drive  
Amherst, NH 03031 USA  
TEL: 603-883-3390