

## CERTIFICATE OF ANALYSIS

Product	100 bp DNA Ladder (Lot No. KI11505)
---------	-------------------------------------

Test Date	2022.04
-----------	---------

### Lot analysis

	Quantity	Agarose check	accredited
100 bp DNA Ladder III	50 µg	V	V

**Range:** 100-3,000 base pairs

**Number of bands :** 12


**Concentration:** 100 µg / ml

**Storage buffer :** 10 mM Tris-HCl (pH8.0), 1.0 mM EDTA

**Recommended loading:** 5 µl (500 ng)

### Test analysis

5 µl of the 100 bp DNA ladder was analyzed on a 1.5% agarose gel.



**IBI Scientific**

**Address:** 9861 Kapp Court | Peosta, IA 52068 **Tel:** 1 800 253 4942 **Fax:** 563 690 0490

**website:** [www.ibisci.com](http://www.ibisci.com) **email:** [info@ibisci.com](mailto:info@ibisci.com)