

# Certificate of Analysis



SCIENTIFIC

800-253-4942 • [IBISCI.com](http://IBISCI.com)

An ISO 9001 Certified Company

## INSTAPAGE - 29

### 30% 29:1 Acrylamide/Bis- Acrylamide Solution

#### PRODUCT NUMBER

IB70003 \_\_\_\_\_ 1L

#### LOT NO.

Lot# \_\_\_\_\_ 22L5004

#### STORAGE

Store at +4 °C. Keep tightly sealed

#### CAS NO.

CAS# \_\_\_\_\_ N/A

#### SPECIFICATIONS

	Specification	Analysis	Disposition
Appearance (Visual Inspection)	PASS / FAIL	Clear, Colorless	PASS
Nucleasae Activity (DNase & RNase, Fluorimetric Analysis)	PASS / FAIL	None Detected	PASS
pH @ 25 C (Acryl 10%, 0.1 M NaCl)	5.5 - 6.5	5.84	PASS
Purity (Acryl)	> 98%	98.4%	PASS
Purity (Bis)	> 99%	99.3%	PASS
Conductivity (µmhos)	> 15	> 15	PASS

#### DATE OF

Manufacture Date: \_\_\_\_\_ 12/29/2022

Restest Date: \_\_\_\_\_ 12/2025

#### CERTIFICATION

We certify that this batch conforms to the specifications listed above and has been reviewed by a quality representative. This document has been electronically produced and signed by IBI SCIENTIFIC.

Abigail Wood  
Quality Control

12/29/22  
Date

FOR RESEARCH AND DEVELOPMENT PURPOSES ONLY

Building Quality Scientific.

IBI SCIENTIFIC • 7445 Chavenelle Road • Dubuque, IA 52002-9664 • USA