








SYMBOL	STANDARD REFERENCE	STANDARD TITLE	SYMBOL TITLE	EXPLANATORY TEXT
	ISO 15223-1 Reference no. 5.1.1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements.	Manufacturer	Indicates the medical device manufacturer, as defined in EU Directive 90/385/EEC, 93/42/EEC and 98/79/EC
	ISO 15223-1 Reference no. 5.3.7	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements.	Temperature limit	Indicates the temperature limits to which the medical device can be safely exposed
	ISO 15223-1 Reference no. 5.4.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements.	Caution	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.
	ISO 15223-1 Reference no. 5.4.2	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements.	Do not re-use	Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure.
	ISO 15223-1 Reference no. 5.1.6	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements.	Catalogue number	Indicates the manufacturer's catalogue number so that the medical device can be identified.
	ISO 15223-1 Reference no. 5.1.5	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements.	Batch code	Indicates the manufacturer's batch code so that the batch or lot can be identified.
	ISO 15223-1 Reference no. 5.1.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements.	Use-by-date	Indicates the date after which the medical device is not to be used.
<b>Rx ONLY</b>	21 CFR 801.109	Labeling; Prescription devices	Prescription use only	Caution: Federal law (USA) restricts this device to sale by or on the order of a licensed healthcare practitioner.