

Specification Sheet

3.5 gram non-sterile Pure Aloe accelerator-free disposable nitrile examination glove

Powder-free, medical grade

- Ambidextrous
- Beaded cuff
- Finger Textured
- Color - Green
- Product Conformance - ASTM D6319
- Complies with FDA Standards CFR Title 21 CFR 117.2600 for food contact regulations

<u>Dimension</u>	<u>Size</u>	<u>Palm Width</u>	<u>Length</u>
	XS	75~5	>240
	S	85~5	>240
	M	95~5	>240
	L	106~5	>240
	XL	115~5	>240

<u>Thickness</u>	<u>Single-Wall (mm)</u>	<u>Typical Value</u>
	Cuff	>0.05 0.08
	Palm	>0.06 0.09
	Finger	>0.08 0.14
	Beading	>.050 0.50

<u>Physical Properties</u>	<u>Un-Aged Spec</u>	<u>Typical Value</u>	<u>Aged Spec</u>	<u>Typical Value</u>
Tensile Strength (MPa)	>14	24	>14	36
Ultimate Elongation (%0	>500	580	>400	600

<u>Quality Specification</u>	<u>Inspection Criteria</u>	<u>Sampling Plan / AQL</u>
	Dimension	S2 - AQL 4.0
	Physical Properties	S2 - AQL 4.0
	Water Tight Test	G1 - AQL 1.0 (Single Normal Sampling)

Country of Origin - Malaysia

