



Powder Free Extra Length Ambidextrous Non-Sterile 26 cm Nitrile Gloves with Textured Fingertips

PPE Category III (Complex Design) according to Council Directive 89/686/EEC

Fully compliant to the latest PPE norms – EN 374:2003 "Protective gloves against chemicals and micro-organisms" **PRODUCT INFORMATION**

Size	Catalogue Numbers	Applicable Norms with Pictograms				
Extra Small (XS/6)	67 6231	67 6231 EN 374:2003				
Small (S/7)	67 6232		Level 3	CE 0120*		
Medium (M/8)	67 6233					
Large (L/9)	67 6234	EN 420:2003 + A1:2009				
Extra Large (XL/10)	67 6235	Also meets or exceeds EN 455-1:2000, EN 455-2:2015, EN 455-3:2015 & EN 455-4:2009 relating to Council Directive 93/42/EEC for Medical Devices				

* SGS UK Ltd (Notified Body No: 0120), Unit 202B Worle Parkway, Weston-super-Mare, BS22 6WA, United Kingdom

Material: Proprietary multi-polymer formulation (Acrylonitrile Butadiene with blend of polychloroprene), based on twinSHIELD™ technology. Contains no natural rubber latex.

Design: Double barrier protection afforded by orange outer layer combined with white inner-lining. Ambidextrous, beaded cuff and with textured fingertips.

Packaging: Ninety (90) gloves per dispenser. Dispenser is varnished for enhanced compatibility with laboratory environments. Gloves are flat-packed. Ten (10) dispensers per case. Packed in a reinforced single-walled shipping case.

PHYSICAL PROPERTIES

Characteristics		Value	Test Method	
	Γ			
Freedom from holes ¹ AQL as defined per ISO 2859 for sam		<0.65 AQL ¹	EN 374:2003	
AQL as defined per 150 2659 for sam	ipting by attributes			
Tensile Properties	Tensile Strength (r	min) Typical	Ultimate Elongation	
- Before Aging	6.0N, min.	7.0N	500%, min.	EN 455-2:2015, ASTM D573-04(2015) and ASTM D412-15a
- After Accelerated Aging	6.0N, min. 8.0N		400%, min.	
Dimensional	MeasuredPoint	Mm	mil	
- Nominal Thickness	Middle Finger	0.17	6.6	ASTM D3767-03(2014)
	Palm	0.14	5.5	
	Cuff	0.10	4.0	
- Length	260mm, min.		265mm, typical	EN 420:2003 + A1:2009

Palm Widths

- Nominal Width (mm)	XS/6	S/7	M/8	L/9	XL/10	EN 455-2:2015
	≤80	85	95	105	≥110	

Hand Circumference						
- Nominal circumference (mm)	XS/6	S/7	M/8	L/9	XL/10	EN 420:2003 +A1:2009
	152	178	203	229	254	

ADDITIONAL DATA

- Biocompatibility demonstrated by Buehler and Primary Skin Irritation Tests.
- Non-detectable levels of chemical allergens using aqueous solution extraction (Phosphate buffered solution) and High Performance Liquid Chromatography (HPLC) assay method for quantitative analysis.
- Accelerator-free to minimize the risk of Allergic Contact Dermatitis (also known as Type IV, Delayed Hypersensitivity or Chemical Allergy).
- **Powder free** to minimize the potential consequences of powder-borne dermatitis. Residual powder content is 1.0 mg/glove (typical) with a limit of 2.0 mg/glove (ISO 21171:2006 "Medical gloves Determination of removable surface powder").
- Micro organism and virus resistant passes highest level of micro organism resistance per EN 374-2:2014 (Performance level 3, AQL <0.65 and inspection level G1 according to 1000ml water test) and passes viral penetration test using Phi-X174 bacteriophage (ISO 16604:2004 Procedure B & ASTM F1671-97b).
- Tested for electrostatic properties according to EN 1149-1/2/3 & 5.
- **Extensively tested for chemical permeation** according to EN 16523-1:2015 (please refer to chemical resistance guide on website www.shieldscientific.com/public/chemical-resistance-guide).

QUALITY SYSTEMS

• Manufactured in accordance with ISO 9001:2015 and ISO 13485:2016.

"SHIELDskin™, A revolution in Glove Technology"



Units 12-13 Lindfield Enterprise Park Lewes Road, Lindfield RH16 2LH Tel: 01444-480444 Fax: 01444-484045 Email: sales@sentinel-laboratories.com Web: www.sentinel-laboratories.com