

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

Product
Produit

Name and address of the applicant
Nom et adresse du demandeur

Name and address of the manufacturer
Nom et adresse du fabricant

Name and address of the factory
Nom et adresse de l'usine

Note: *When more than one factory, please report on page 2*
Note: *Lorsque il y a plus d'une usine, veuillez utiliser la 2^{eme} page*

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Trademark (if any)
Marque de fabrique (si elle existe)

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

Model / Type Ref.
Ref. De type

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{eme} page

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

CERTIFICAT D'ESSAI OC

Switching Power Supply

MEAN WELL ENTERPRISES CO LTD
28 WUQUAN 3RD RD WUGU DIST
NEW TAIPEI, 248 Taiwan

MEAN WELL ENTERPRISES CO LTD
28 WUQUAN 3RD RD WUGU DIST
NEW TAIPEI, 248 Taiwan

MEAN WELL ENTERPRISES CO., LTD.
NO.28, WUQUAN 3RD RD., WUGU DIST., NEW TAIPEI CITY
248
TAIWAN (R.O.C.)

Additional Information on page 2
See Page 2

MW, MEAN WELL



MSP-100-x
See Page 2

Additional Information on page 2

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

E227340-D10-CB-2 issued on 2015-11-13

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2015-11-18

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-49915-UL

Model Details:

MSP-100-x (x can be 3.3, 5, 7.5, 12, 15, 24, 36 or 48; x is to denote different output voltage)

Factories:

SUZHOU MEAN WELL TECHNOLOGY CO., LTD.

NO. 77, JIAN-MIN ROAD, DONG-QIAO, PAN-YANG IND. PARK, HUANG-DAI TOWN, XIANG-CHENG DISTRICT, SUZHOU, JIANGSU 215152

P. R. CHINA

MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD.

2F, A BUILDING, YUEAN INDUSTRIAL PARK, HUANGCUN, DONGPU TOWN, TIANHE DISTRICT, GUANGZHOU, GUANGDONG

P.R. CHINA

Ratings:

Input: 100-240 V ac, 50/60 Hz, 1.4-0.6A

Output:

Model name	Output rating	Output power
MSP-100-3.3	3.3Vdc, 20A	66W
MSP-100-5	5Vdc, 17A	85W
MSP-100-7.5	7.5Vdc, 13.5A	101.3W
MSP-100-12	12Vdc, 8.5A	102W
MSP-100-15	15Vdc, 7A	105W
MSP-100-24	24Vdc, 4.5A	108W
MSP-100-36	36Vdc, 2.9A	104.4W
MSP-100-48	48Vdc, 2.2A	105.6W

Additional Information:

Additionally evaluated to EN 60601-1:2006/ A1:2013/ A12:2014. National Differences specified in the CB Test Report. Risk management was not included in this investigation.

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2015-11-18

Signature:

Jan-Erik Storgaard