

Well Water – Test Kit

The **Well Water – Test Kit** was specifically designed to help you test quickly and easily for the most common contaminants found in private well water, including: iron, copper, lead, bacteria, pesticides, nitrates, nitrites, chlorine, pH and hardness.

If your home draws from private well water, regular testing (recommended yearly or more) is the only way to ensure the purity of your water and safeguard the health of your family.

What results are expected?

The **Well Water – Test Kit** gives information on the following parameters and their content in tested water. EPA maximum contaminant levels or guideline standards:

Bacteria	None
Lead	Below 15 ppb
Pesticides	Below 3 ppb (atrazine), Below 4ppb (simazine)
Copper	Below 1.3 ppm
Iron	Below 0.3 ppm
Nitrite	Below 1.0 ppm
Total Nitrate/Nitrite	Below 10.0 ppm
Total Chlorine	Below 4 ppm
pH	6.5 to 8.5
Total Hardness	50 ppm or less

Before you begin...

1. Each **Well Water – Test Kit** contains:
 - One Bacteria Test Vial
 - One Lead / Pesticide Test Packet
 - One Copper Test Packet
 - One Iron Test Packet
 - One Nitrate / Nitrite Test Packet
 - One pH / Hardness / Chlorine Test Packet
2. Do not open packets or vial until you are ready to perform the tests.
3. Keep tests out of reach of children or pets.
4. Do not ingest anything from this test kit. Do not drink water sample used for testing.
5. Do not use on hot water or water containing bleach detergents. Do not re-use any part of the test kit.
6. Store and use at room temperature (10-30 °C).
7. Read and follow all instructions carefully.

Please note that this is a screening test and can not be used to certify water as safe or unsafe for drinking. It provides approximate results ONLY when used in strict accordance with instructions.

Test Instructions

Bacteria Test

1. Take out the Bacteria Test vial and set upright on a flat surface.
2. Collect water sample or turn on tap to a very slow stream.
3. Carefully twist off cap and fill vial to 1/2 inch below the top (to 5 ml line). Do not overfill and do not spill the bacterial growth powder in the vial.
4. Replace cap and twist on tightly. Shake the vial vigorously for 20 seconds.
5. Place the capped vial upright in a warm area (20-30 °C) where it cannot be disturbed for 48 hours.
6. After 48 hours, observe the color of the liquid without opening the vial:

Purple Color: Negative Result (No bacteria were detected).



Yellow Color: Positive Result (It is highly likely that potentially harmful bacteria were detected).



7. For positive results, add bleach to the sample before pouring down the toilet, then wash hands carefully. Negative samples may be poured directly into the toilet. Discard vial in the trash.


Lead / Pesticide Test

The **Lead Test** can detect dissolved lead at levels below the EPA Action level of 15 parts per billion (ppb). The **Pesticide Test** detects two of the most common pesticides at or below the EPA Maximum Contaminant Level (atrazine - 3 ppb, and simazine - 4 ppb).


1. Open **Lead / Pesticide** packet and take out all contents. The packet contains a test vial, a dropper pipette, two test strips, and a desiccant (to be discarded).

- Using dropper, place **exactly two dropper-fuls** of water sample into test vial. To pick up the sample, tightly squeeze the bulb at the end of the dropper and place the open end into the water sample. Release the bulb to pick up the sample, then squeeze again to expel the sample into the vial.
 - Swirl vial gently for several seconds. Place on a flat surface.
 - Place both test strips into the test vial, with arrows pointing DOWN.
 - Wait 10 minutes. Do not disturb strips or vial during this time. Blue lines will appear on the strips.
 - Take the strips out of the vial and lay them on a flat surface with the arrows pointing to the LEFT.
- Read Results.

Negative: If you only see the LEFT line (next to number 11), or the LEFT line (next to number 11) is darker than the RIGHT line (next to number 22).



Positive: The RIGHT line (next to number 22) is darker than the LEFT line (next to number 11), or lines are equally dark (Both LEFT and RIGHT lines are equally dark)



- Note: If no lines appear, or both lines are very light, the test did not run properly and the result is not valid.

If a test strip shows a positive result, your water sample may contain lead or pesticides at a toxic level.

Copper Test

- Open Watersafe Copper Test kit and remove the Copper Test strip from the foil packet.
- Immerse the reagent pad into water sample for 30 seconds, waving in a gentle back and forth motion.
- Remove the strip and shake once to remove excess water.
- Wait 2 minutes, then match result to the color chart below.

Kupfer




Iron Test

- Open **Iron** packet and take out test strip.
- Immerse the reagent pad into water sample for 5 seconds, waving in a gentle back and forth motion.
- Remove the strip and shake once to remove excess water.
- Wait 2 minutes, then match result to the color chart below.



Iron



Nitrate / Nitrite Test

1. Carefully open **Nitrate / Nitrite Test** packet and take out test strip.
2. Immerse the reagent pads into water sample for 2 seconds, remove, after 1 minute match colors with the chart below.
3. Colors are stable for 1 minute.

Total Nitrate / Nitrite (end pad)



Nitrite (pad nearest handle)



pH / Hardness / Chlorine Test

1. Carefully open **pH / Hardness / Chlorine Test** packet and take out strip.
2. Immerse the reagent pads into water sample and remove immediately. Hold the strip level for 15 seconds.
3. Match pH, Total Hardness and Total Chlorine pads (in that order) to the color chart.

pH (end pad)



LR Total Hardness (middle pad)



Total Chlorine (pad nearest handle)

