



# 1 Test Results

EUT Name	CPAP System		
EUT Model	RXiBreeze 20A Plus	EUT No.	GB-H2001989
Applicant	Resvent Medical Technology Co.,Ltd	Address	Room-602,Building B&C,Gaoxinqi Industrial Park,Liuxian No.1 Road,Xingdong Community,Bao'an,518100 Shenzhen,PEOPLE'S REPUBLIC OF CHINA
Manufacturer	Resvent Medical Technology Co.,Ltd	Address	Room-602,Building B&C,Gaoxinqi Industrial Park,Liuxian No.1 Road,Xingdong Community,Bao'an,518100 Shenzhen,PEOPLE'S REPUBLIC OF CHINA
Received Date	2019-09-17	Test Date	2019-09-18
Client Test Participant	Hao Tang、 Alice Chen		
Test address	No.163, Pingyun Road, West of Huangpu Avenue, Guangzhou. Guangdong.China		
Test Standard	1. RTCA/DO-160G: 2010 <i>Environmental Conditions and Test Procedures for Airborne Equipment</i> 2. The test condition and working state of the sample are confirmed by the E201909048946-1electromagnetic compatibility testing status form of the customer.		

## Test Results

No.	Test Items	Equipment Category	Test Results	Report Sections	Comment
1	Section 21 Radiated RF Emission(RE)	Category M	Pass	3.1	100MHz~6GHz

**Test Result: PASS**

Prepared by Lingzhou Lin

Reviewed by Bruce Li

Approved by Yong Dai



Position Manager

Test Location: EMC Laboratory of Guangzhou GRG Metrology and Test Co., Ltd. (Special Seal)

Issue Date: 2019.9.25