

Model LM-8000

4-in-1 Pocket Thermo-Anemometer, Hygrometer, Thermometer & Illuminometer

Instruction Manual





REED

822

LM-8000

Table of Contents

Features	2-3
Specifications	3-4
Instrument Description	5
Measuring Procedures	5-6
Air Velocity Measurement	5
Temperature Measurement (Thermocouple)	5-6
Relative Humidity Measurement	6
Light Measurements	6
Additional Functions	7
Hold Function	7
Data Record Function	7
Auto Power Off	7
Battery Replacement	8

For service on this or any other REED product or information on other REED products, contact REED Instruments at info@reedinstruments.com



Features

- 4-in-1 multi-function meter... anemometer, humidity meter, light meter and type-K digital thermometer
- Type K (NiCr-NiAl) thermocouple input jack
- High precision thin-film capacitance humidity sensor
- Photo diode and colour correction filter light sensor
- Velocity Units: m/s, kph, mph, knots, and fpm
- °F or °C and Lux or Ft-cd selectable
- Data Hold, Record (Max/Min)
- Includes batteries

Specifications

Power Supply:	Single 9V battery
Dimensions:	6.1 x 2.4 x 1.3" (156 x 60 x 33mm)
Weight:	8.8 oz (160g)
Optional Accessories:	Air/Gas Probe (LS-103) Right Angle Surface Probe (LS-104) General Purpose Probe (LS-107) Surface Probe (LS-109) Needle Tip Probe (LS-134A) Surface Probe (LS-139) Soft Carrying Case (CA-52A)

		Range	Resolution
Air velocity	ft/min	80 to 5910 ft/min	1 ft/min
	m/s	0.4 to 30.0 m/s	0.1 m/s
	km/h	1.4 to 108.0 km/h	0.1 km/h
	MPH	0.9 to 67.0 mile/h	0.1 MPH
	knots	0.8 to 58.3 knots	0.1 knots
	Temperature	32 to 122°F	0.1°F
	(thermister)	0 to 50°C	0.1°C



		Range	Resolution	
	% RH	10 to 95 %RH	0.1 %RH	
Humidity	Temperature	32 to 122°F	0.1°F	
	(thermister)	0 to 50°C	0.1°C	
Light *auto range	Lux	0 to 2,200 Lux	1 Lux	
	Lux	1,800 to 20,000 Lux	10 Lux	
	Ft-cd	0 to 204.0 Ft-cd	0.1 Ft-cd	
	FI-CO	170 to 2,000 Ft-cd	1 Ft-cd	
Temperature (Type K)		-148 to 2372°F	0.1°F	
		-100 to 1300°C	0.1°C	

	Range	Accuracy			
Air velocity	80 to 5910 ft/min				
	0.4 to 30.0 m/s	≤ 20 m/s : ± 3% F.S.			
	1.4 to 108.0 km/h	> 20 m/s : ± 4% F.S.			
	0.9 to 67.0 mile/h	> 20 m/3 . ± 470 1.0.			
	0.8 to 58.3 knots				
	32 to 122°F	± 2.5°F			
	0 to 50°C	± 1.2°C			
	10 to 95 %RH	< 70% RH : ± 4 %RH; ≥70% RH : ± (4%rdg + 1.2 %RH)			
Humidity	32 to 122°F	± 2.5°F			
	0 to 50°C	± 1.2°C			
Light	0 to 20,000 Lux	± 5% rdg ± 8 dgt			
Light	0 to 2,000 Ft-cd				
Temperature	-148 to 2372°F	± (1% rdg + 2°F)			
(Type K)	-100 to 1300°C	± (1% rdg + 1°C)			

ft/min = feet per minute m/s = meters per second km/h = kilometers per hour MPH = miles per hour

knots = nautical miles per hour

Ft-cd = foot candle



Instrument Description

- 1. Power Button
- 2. Hold Button
- 3. Max/Min Button
- 4. Unit/Zero Button
- 5. °C/°F Button & Lux/Ft-cd Button
- 6. Function Button
- 7. Air Flow Sensor
- 8. Thermocouple Input Socket
- 9. R.H. Sensor
- 10. Light Sensor
- 11. LCD display
- 12. Battery Compartment/ Cover
- 13. Wristlet



Operating Instructions

Air Velocity Measurement

- 1. Power on the instrument by pressing the Power Button
- 2. Select the Anemometer function by pressing Function Button
- 3. Press the Unit/Zero Button to select unit that you want and then face the Air Flow Sensor to the source of wind
- 4. Allow time for the reading to become stable and note the value indicated



Thermocouple Temperature Measurement

- 1. Power on the instrument by pressing the Power Button
- 2. Plug a type K thermocouple probe in the Thermocouple Input Socket
- 3. Select the Temperature function by pressing Function Button
- Contact the Thermocouple Sensor Head with the object being measured and the reading value will be displayed on the LCD display

Humidity & Ambient Temperature Measurement

When the meter is applied in a new environment, it will take a few minutes to reach a stable condition

- 1. Power on the instrument by pressing the Power Button
- 2. Select the Relative Humidity function by pressing the Function Button
- 3. The reading value of relative humidity and temperature will be displayed on the LCD display

Light Measurement

- 1. Power on the instrument by pressing the Power Button
- 2. Select the Light Measurement function by pressing the Function Button until the light value is displayed
- 3. Press the Lux/Ft-cd Button to select measuring unit Lux or Ft-cd

Zero Offset Adjustment of Light Function

- For best results you should zero the light sensor prior to use in a dark environment. To achieve this, place the light sensor end of the meter under a desktop or flat surface to block any light. Press the Unit/Zero Button setting the meter indication to zero.
- The zero point can drift due to environment temperatures, battery power, as well as various other reasons. It is recommended that the zero be checked frequently using the above procedure.



Hold Function

Pressing the Hold Button will freeze the current value with a HOLD symbol on the display. Press again to release the hold function.

Data Record Function

The Data Record function records & displays the maximum and minimum reading values. Start the Data Record function by pressing the Max./Min. Button once. There will be a "REC" symbol on the display. With the REC symbol on the display:

- 1. Press the Max/Min Button once and the Max symbol along with the maximum value will appear on the display.
- 2. Press the Max/Min Button again, the "Min" symbol along with the minimum value will appear on the display.
- To exit the memory record function, press the Max/Min Button continuously for at least 2 seconds. The display will revert to the current reading.
- 4. To clear the Max/Min value recorded by pressing the Hold Button once. The previously recorded Max/Min value will be lost and then revert to the REC function continuing to record.

Auto Power Off

In order to prolong the battery life, the instrument has "Auto Power Off" function. The meter will switch off automatically if no buttons are pressed for around 10 minutes.



Battery Replacement

When the left corner of the LCD display shows the low battery symbol, this indicates that replacement of the battery is then needed. However specification measurements could still be taken for another few hours before the meter becomes inaccurate.

- 1. Open the Battery Cover at the back of the meter and remove the battery
- 2. Replace with a 9V battery (alkaline or heavy-duty type) and reinstall the cover

For service on this or any other REED product or information on other REED products, contact REED Instruments at info@reedinstruments.com

Notes _				
				_
				_

