



Novitatech  
Technology division of  
Novita Disability Services  
ABN 53 137 537 636

Issued to: Push Mobility

Push Mobility  
341 George Street  
Fitzroy 3065 Vic

171 Days Road  
Regency Park SA 5010  
PO Box 2438  
Regency Park SA 5942

T: 1300 85 55 85  
E: [testing@novita.org.au](mailto:testing@novita.org.au)

[www.novitatech.org.au](http://www.novitatech.org.au)

Novitatech Test Laboratory  
NATA Accreditation Number 2953

# Certificate of Compliance

**Product Details:** Push Mobility Beach Wheelchair

**Product Name:** Mobi-Chair floating Beach Wheelchair

The product listed above complies with the requirements of AS/NZS 3695.1:2011 (Excluding sections or clauses in the report marked not tested or not applicable due to the wheelchair being a specialised product)

for user weight 120 kg

Yours sincerely

A handwritten signature in black ink, appearing to read 'Wayne Wurfel'.

Wayne Wurfel  
Senior Test Technician  
Authorised Signatory  
[testing@novita.org.au](mailto:testing@novita.org.au)

Date of Issue 21/09/2018

