



## CERTIFICATE OF CONFORMANCE

### ISOBUTANE

#### INSTRUMENT PURITY

<u>Guaranteed Specifications</u>		<u>Minimal Purity</u>
99.5 % +	Isobutane	99.50 % +
Sum of all impurities to be < 0.50 %	Methane	< 1 PPM
	Ethane	<1 PPM
	Propane	< 70 PPM
Water < 5 PPM	n-Butane	< 2000 PPM
	Isopentane	< 2 PPM
Liquid Phase Analysis	Water	< 3 PPM
	Sulfur	Non Detected
	Nitrogen	< 10 PPM
	Oxygen	< 5 PPM

Certified by: \_\_\_\_\_

Date: \_\_\_\_\_

SO# 00093096

Cylinder Serial Number(s)

Batch Number  
IBU-092717-I