

CERTIFICATE OF CONFORMANCE

PROPANE

INSTRUMENT PURITY

Guaranteed Specifications

Minimal Purity

99.5 % +

Propane

99.50% +

Sum of all impurities to

Methane

< 1 PPM

be < 0.50 %

Ethane

< 140 PPM

Water < 5 PPM

Propylene

< 10 PPM

Isobutane

< 2200 PPM

n-Butane

< 1500 PPM

Isopentane

< 3 PPM

Carbon Dioxide

< 1 PPM

Liquid Phase Analysis

Water

< 5 PPM

Sulfur

Non Detected

Certified by: _____

Date: _____

Cylinder Serial Number(s)

SO# 00093096

Batch Number

PRO-092717-I