



# CALIBRATION REPORT

## PI TAPE®

Report Number:	
Date Issued:	
Tape Serial Number:	
Tape Size:	12" – 24" O.D.

### About this Calibration

Temperature	68 Degrees F ± 2 Degrees	
Relative Humidity	45% ± 15%	
Tension	Five Pounds	
Tape Tolerance	Within .001"	Within Mfg Tolerance of .001"
Uncertainty of Measurement	±.00025 Inches	

Calibration Performed by: Daniel Brinkmeier

### Actual Findings and Procedure Numbers

Tape Reads		Over 20.0000"	Procedure 1
------------	--	---------------	-------------

Remarks: New Gage

### Calibration Equipment Used

Gage	Serial Number	Test Date	Due Date	NIST Traceable Report Number
20"	042199	Jan-12-16	Jul-12-16	3370680001

This tape has been checked at one or more points over a calibrated measuring standard. Calibration performed in accordance with ANSI/NC SL Z540-1-1994, ISO/IEC 17025:2005 and the former MIL-STD 45662A.

**All standards are traceable to the National Institute of Standards and Technology**

This calibration report shall not be reproduced, except in full, without the written approval of the issuing laboratory. Results listed relate only to the item being calibrated.



**Pi Tape Texas, LLC.**

10235 Robinson Drive, Tyler, TX 75703  
 Phone (903) 266-9204 Fax (903) 939-0909  
[www.pitape.com](http://www.pitape.com)