



NUTRIACTIVA LLC DBA BETA TECHNOLOGY

10800 NESBITT AVE S, BLOOMINGTON, MN 55437

NUTRIACTIVA.COM | INFO@NUTRIACTIVA.COM

CALIBRATION CERTIFICATION

OWNER: ABC Company, Phoenix AZ

DATE: January 16, 2024

PRODUCT: Lange Skinfold Caliper

DATE OF PURCHASE: September 27, 2023

AGE OF CALIPER: Less than 1 year

SERIAL NUMBER: 1383-2231

DIAL CALIBRATION

Factory Standards are +/- 1.0mm

INCREMENT	DIAL READING	DIFFERENCE
10mm	10mm	0
20mm	20mm	0
30mm	30.25mm	0.25mm
40mm	40mm	0
50mm	50mm	0

Calibrated Not Calibrated

PRESSURE CALIBRATION

Spring should be replaced every 5 years or after 10,000 uses, whichever comes first.

Calibrated Not Calibrated

Certified By

Joseph Richards

(866) 731-8660