

120SC ADVANCED VACCINE FRIDGE THERMOMETER

FEATURES:

- · Dual channel measurement
- · Low/high alarm limit setting
- · Minimum and maximum reading memories with real time and date stamp
- · Flashing light alarm indicator
- °C/°F switchable
- · Time and date display
- · Low battery indicator
- Flip-out desk stand
- Magnetic attachment





OI LOII IOATION	
Measuring Range	-50°C to 70°C (-58°F to 158°F)
Display Accuracy	± 1°C
Display Resolution	0.1°
Display Reading Update	10 seconds
Alarm Limit Setting Resolution	1°
Time Accuracy	± 3 seconds per day
Time Display Format	24 hours
Date Display Format	Day/month/year
Effective Calendar Period	2011-2099
Battery	1.5 volt, type AAA x 2 pieces
Probe cord length	1,000 mm
Display Size	90(W) x 16(H) mm
Product Size	135(W) x 45(H) x 23(D) mm
Accessories	1. 1.5 volt, type AAA battery x 2 pieces 2. Magnet on back of unit 3. Sensor holder with double sided tape x 2 pieces

INSTALLATION

- 1. Open battery cover.
- 2. Install batteries with correct polarity positioning.
- 3. Close battery cover.
- 4. Fully extend the probe wires.
 - a) If using glycol tubes (available separately) then place glycol tubes in a vaccine box inside fridge.
 - b) If not using glycol tubes, please place sensors in a secure area inside the fridge away from the wall on the fridge.
- 6. Peel off LCD display protective sheet.

OPERATION

Select Temperature Unit

Slide the ${^{\circ}}\dot{C}/{^{\circ}}F$ switch on the back to select temperature unit.

Time And Date Display

Press TIME/DATE to switch the display to show the current time or date.

In date display mode, the display will automatically switch back to time display if there is no operation for 10 seconds.

Time Setting

- 1. In time display mode, press and hold SET for about 3 seconds to enter time setting and the hour (HR) digits will be flashing.
- 2. Press ▲ to adjust the digits to the current hour.
- 3. Press SET once to confirm the hour setting and then the minute (MIN) digits will begin to flash.
- 4. Press to adjust the digits to the current minute(s).
- 5. Press SET once to confirm the minute setting and then the second (SEC) digits will begin to flash.
- 6. Press ▲ to adjust the digits to the current second.
- 7. Press SET again to finish time setting and return to normal display.

Date Setting

- 1. In date display mode, press and hold SET for about 3 seconds to enter date setting and the Year digits will be flashing.
- 2. Press lacktriangle to adjust the digits to the last two digits of the current year.
- 3. Press **SET** once to confirm the year setting and then the month digits will begin to flash.
- 4. Press ▲ to adjust the digits to the current month.

- 5. Press **SET** once to confirm the month setting and then the date digits will begin to flash.
- 6. Press to adjust the digits to the current date.
- 7. Press **SET** again to finish date setting and return to normal display.

Minimum / Maximum Reading Memory

- 1. Press MODE once to display the minimum measured reading and the time display will show when this minimum reading was recorded.
- 2. Press **TIME/DATE** to see the date of when the minimum reading was recorded.
- 3. Press MODE once to display the maximum measured reading and the time display will show when this maximum reading was recorded.
- 4. Press TIME/DATE to see the date of when the maximum reading was recorded.
- 5. Press **MODE** three times to return to normal display.
- 6. To clear the minimum and maximum reading memories, press **CLEAR** once when the minimum or maximum readings are displayed. All readings for both channels and time stamps will be reset to current readings and time.
- 7. Clear the memory once before taking new minimum or maximum reading.

Low / High Alarm Setting

- 1. Press MODE three times to enter low / high alarm setting and the LO icon will be displayed.
- 2. Press and hold SET for about 3 seconds and the LO digits will be flashing.
- 3. Press to input the desired low temperature limit.
- 4. Press SET to confirm low temperature alarm limit. The HI icon will be displayed and the digits will begin to flash.
- 5. Press **\(\rightarrow\)** to input the desired high temperature limit.
- 6. Press SET to confirm the high temperature alarm limit and finish the setting.
- 7. The alarm will sound when the temperature reading is lower or higher than the alarm limit set. Press any button to stop the alarm sound but the HI or LO icon will still flash until the reading is within the set limits.

NOTE: If connectors are unplugged, the alarm will go off. Once re-connected, the alarm will stop.

Alarm ON/OFF

Press 🔔 to switch alarm OFF or ON.

Alarm Indicator

When the reading exceeds the alarm limit the red light will be on and continue to flash to indicate that an alarm has been triggered. Press and hold **CLEAR** for about 3 seconds to cancel the flashing red light.

Switch Channel Display

Press \$1/\$2 once to switch the display for channel \$1 or \$2 reading.

NOTE

- Do not operate the thermometer in an environmental temperature lower than 0°C/32°F or higher than 50°C/122°F otherwise incorrect readings or damage to the thermometer may result.
- 2. If the thermometer is not in use for a long period of time then remove the batteries from battery compartment to avoid battery leakage.

ERROR SYMBOLS

SYMBOL	DESCRIPTION	ACTION REQUIRED
	Low battery voltage	Replace the batteries
LLL	Sensor open circuit The reading is out of low range (-50°C) Connector is unplugged	Return the thermometer for repair Keep the measurement above low range Ensure the connectors are properly inserted
ННН	Sensor short circuit The reading is out of high range (70°C)	Return the thermometer for repair Keep the measurement below high range

ONE YEAR WARRANTY

If this product proves to be defective in material or workmanship within one year of purchase, please return it to the address below. It will be repaired or replaced without charge upon receipt of the unit prepaid with \$5.00 to cover handling, packaging and return postage. Please include proof of purchase, your full name, address, daytime phone number or email address.

