

AARON 1250™



User's Guide

Bovie

Bovie

AARON 1250™



USER'S GUIDE

This manual and the equipment it describes are for use only by qualified medical professionals trained in the particular technique and surgical procedure to be performed. It is intended as a guide for using the Aaron 1250™ Electrosurgical Generator only.

Additional technical information is available in the Aaron 1250™ Electrosurgical Generator Service Manual. For the latest user information and technical bulletins, visit www.boviemed.com.

Equipment Covered in this Manual

Aaron 1250 Electrosurgical Generator:
Model No.: A1250U

For Information Call

Bovie Medical Corporation • 5115 Ulmerton Road, Clearwater, Florida 33760 USA
U.S. Phone 1-800-537-2790 Fax 1-800-323-1640 • International Phone +1-727-384-2323 Fax +1-727-347-9144
www.boviemed.com • sales@boviemed.com



Emergo Europe
Molenstraat 15
2513 BH, The Hague
The Netherlands



©2015 Bovie Medical Corporation
All rights reserved.

Contents of this publication may not be reproduced without the written permission of Bovie Medical Corporation.

Part number, MC-55-128-001 Rev. 8

CONVENTIONS USED IN THIS GUIDE

WARNING

Indicates a potentially hazardous situation which, if not avoided, could result in death or serious injury.

CAUTION

Indicates a hazardous situation which, if not avoided, may result in minor or moderate injury.

NOTICE

Indicates an operating tip, a maintenance suggestion, or a hazard that may result in product damage.

TABLE OF CONTENTS

Equipment Covered in this Manual	iii
For Information Call	iii
Conventions Used in this Guide	iii
Introducing the AARON 1250™ Electrosurgical Generator	1-1
Indications for Use	1-2
Operating Principles	1-2
Intended Use	1-2
Key Features	1-2
Components and Accessories	1-3
Additional Accessories	1-3
Safety	1-3
Controls, Indicators, and Receptacles	2-1
Front Panel	2-2
Symbols on the Front Panel	2-3
Cut and Blend Controls	2-4
Coag and Bipolar Controls	2-5
Indicators	2-6
Power Switch and Receptacles	2-7
Rear Panel	2-8
Symbols on the Rear Panel	2-8
Getting Started	3-1
Initial Inspection	3-2
Installation	3-2
Function Checks	3-2
Setting Up the Unit	3-2
Checking the Return Electrode Alarm	3-2
Confirming Modes	3-3
Checking Bipolar Mode (with footswitch)	3-3
Checking Monopolar Mode (with footswitch)	3-3
Checking Monopolar Mode (with handswitch)	3-3
Performance Checks	3-3
Using the Aaron 1250™	4-1
Inspecting the Generator and Accessories	4-2
Setup Safety	4-2
Setting Up	4-4
Preparing for Monopolar Surgery	4-5
Applying the Return Electrode	4-5
Connecting Accessories	4-5
Preparing for Bipolar Surgery	4-5
Activation Safety	4-6
Activating the Unit	4-7
Maintaining the Aaron 1250™	5-1
Cleaning	5-2
Periodic Inspection	5-2
Fuse Replacement	5-2
Troubleshooting	6-1

Repair Policy and Procedures	7-1
Responsibility of the Manufacturer	7-2
Returning the Generator for Service.....	7-2
Step 1 – Obtain a Returned Goods Authorization Number	7-2
Step 2 – Clean the Generator.....	7-2
Step 3 – Ship the Generator.....	7-2
Technical Specifications	A-1
Performance Characteristics	A-2
Input Power	A-2
Duty Cycle	A-2
Dimensions and Weight	A-2
Operating Parameters	A-2
Transport	A-2
Storage	A-2
Audio Volume	A-3
Return Electrode Sensing	A-3
Low Frequency (50-60 Hz) Leakage Current.....	A-3
High Frequency (RF) Leakage Current.....	A-4
Standards and IEC Classifications.....	A-4
Operating Conditions.....	A-4
Class I Equipment (IEC 60601-1)	A-4
Type BF Equipment (IEC 60601-1) / Defibrillator Proof.....	A-4
Drip Proof (IEC 60601-2-2)	A-4
Electromagnetic Interference.....	A-4
Electromagnetic Compatibility (IEC 60601-1-2 and IEC 60601-2-2)	A-4
Voltage Transients (Emergency Generator Mains Transfer).....	A-4
EMC Compliance	A-4
Output Characteristics.....	A-8
Maximum Output for Monopolar and Bipolar Modes	A-8
Output Power Curves	A-9
Warranty	B-1

LIST OF FIGURES

Figure 2 – 1 Layout of controls, indicators, and receptacles on the front panel2-2
Figure 2 – 2 Controls for the Cut and Blend Modes2-4
Figure 2 – 3 Controls for the Coagulation, Fulguration, and Bipolar Modes.....2-5
Figure 2 – 4 Indicators for power, return electrodes, and footswitch control2-6
Figure 2 – 5 Location of the unit power switch and front panel receptacles.....2-7
Figure 2 – 6 Layout of connectors and controls on the rear panel2-8
Figure 5 – 1 Fuse holder5-2
Figure A – 1 Output voltage (V_{peak}) versus power setting (Cut, Coag)A-9
Figure A – 2 Output voltage (V_{peak}) versus power setting (Bipolar)A-9
Figure A – 3 Output power versus power setting for all modes.....A-10
Figure A – 4 Output power vs impedance for Cut modeA-10
Figure A – 5 Output power vs impedance for Blend modeA-11
Figure A – 6 Output power versus impedance for Coagulation modeA-11
Figure A – 7 Output power versus impedance for Fulguration modeA-12
Figure A – 8 Output power vs impedance for Bipolar mode.....A-12

INTRODUCING THE AARON 1250™ ELECTROSURGICAL GENERATOR

This section includes the following information:

- Indications for Use
- Operating Principles
- Intended Use
- Key Features
- Components and Accessories
- Safety
- Contraindications
- Application Specification

CAUTIONS

Read all warnings, cautions, and instructions provided with this generator before using.

Read the instructions, warnings, and cautions provided with electrosurgical accessories before using. Specific instructions are not included in this manual.

INDICATIONS FOR USE

The Aaron 1250™ Electrosurgical Generator is used to deliver RF energy via an assortment of surgical devices to cut and coagulate different kinds of tissue. For the latest user information and technical bulletins, visit www.boviemed.com.

OPERATING PRINCIPLES

The Aaron 1250™ Generator is a high frequency isolated generator featuring cutting up to 120 watts, a blend mode, 2 coagulation modes and 1 bipolar mode. The generator offers a monopolar handpiece output, monopolar foot controlled output and bipolar foot controlled output. The generator has a return electrode contact and quality monitoring system (NEM) to reduce the risk of patient burns at the return electrode site. The pad-sensing feature allows the user to use either a split or solid return electrode.

INTENDED USE

The Aaron 1250™ Electrosurgical Generator is intended for cutting, coagulation, ablation of tissue in general, gynecologic, orthopedic, ENT and urological procedures performed in an operating suite and procedure room.

NOTICE:

The Aaron 1250™ is not intended for Tubal Ligation.

KEY FEATURES

The Aaron 1250™ Electrosurgical Generator includes the latest technology. This unit offers unsurpassed performance, flexibility, reliability, and convenience.

It includes the following features:

- Two levels of coagulation: Pinpoint Coagulation and Fulguration
Pinpoint Coagulation provides precise control of bleeding in localized areas.
Fulguration provides greater control of bleeding in highly vascular tissue over broad surface areas.
- Return electrode sensing and contact quality monitoring
The Aaron 1250™ incorporates a return electrode contact quality monitoring system (Bovie NEM™). This system determines the type of patient return electrode: single or split plate. The system also continually monitors the contact quality between the patient and the split-plate return electrode. This feature is designed to eliminate patient burns at the return electrode site.

NOTICES:

The Bovie NEM™ system recommends that you use a split-plate patient return electrode.

Before activation, pad placement and visual verification of the split return electrode (split pad) indicator on the front panel is recommended. After connecting the split pad to the generator and placing the split pad securely to the patient, give the unit 5 to 10 seconds to recognize the split pad. The split pad indicator will illuminate green. If the split pad and cord are attached to the generator without secure contact to the patient, the alarm indicator will illuminate red.

- Memory
The unit automatically powers up to the Cut and Coag modes and their last selected power settings.
- Isolated RF output
This minimizes the potential of alternate site burns.
- Standard connectors
These connectors accept the latest monopolar and bipolar instruments.

- Self diagnostics

These diagnostics continually monitor the unit to ensure proper performance.

COMPONENTS AND ACCESSORIES

To avoid incompatibility and unsafe operation, we recommend using the following Bovie® brand accessories supplied with your generator:

- Aaron 1250™ Electrosurgical Generator
- Hospital-grade power cords - 09-039-001; 09-035-001
- User's Guide (s)
- One disposable pencil - ESP1-S
- Three electrodes - ES20 (ball); ES02 (needle); ES01 (blade)
- One reusable grounding cord - A1252C
- Five disposable split grounding pads -ESRE-1
- ESU Series I DVD

ADDITIONAL ACCESSORIES

To avoid incompatibility and unsafe operation, we recommend using the following Bovie accessories with the A1250™:

- BV-1253B - Footswitch for Monopolar and Bipolar procedures

SAFETY

The safe and effective use of electrosurgery depends to a large degree on factors solely under the control of the operator. There is no substitute for a properly trained and vigilant medical staff. It is important that they read, understand, and follow the operating instructions supplied with this electrosurgical equipment.

Physicians have used electrosurgical equipment safely in numerous procedures. Before starting any surgical procedure, the surgeon should be familiar with the medical literature, complications, and hazards of using electrosurgery in that procedure.

To promote the safe use of the Aaron A1250™ Electrosurgical Generator, this section presents the warnings and cautions that appear throughout this user's guide. So that you can operate this equipment with maximum safety, it is important that you read, understand, and follow the instructions in these warnings and cautions. It is also important that you read, understand, and follow the instructions for use in this user's guide.

WARNINGS

Hazardous Electrical Output - This equipment is for use only by trained, licensed physicians.

Danger: Fire / Explosion Hazard - Do not use the Aaron 1250™ electrosurgical generator in the presence of flammable anesthetics.

Fire / Explosion Hazard - The following substances will contribute to increased fire and explosion hazards in the operating room:

- Flammable substances (such as alcohol based skin prepping agents and tinctures)
- Naturally occurring flammable gases which may accumulate in body cavities such as the bowel
- Oxygen enriched atmospheres
- Oxidizing agents (such as nitrous oxide [N₂O] atmospheres).

The sparking and heating associated with electrosurgery can provide an ignition source. Observe fire precautions at all times. When using electrosurgery in the same room with any of these substances or gases, prevent their accumulation or pooling under surgical drapes, or within the area where electrosurgery is performed.

The use of flammable anesthetics or oxidizing gases such as nitrous oxide (N₂O) and oxygen should be avoided if a surgical procedure is carried out in the region of the thorax or the head, unless these agents are sucked away.

Non-flammable agents should be used for cleaning and disinfection wherever possible.

Flammable agents used for cleaning or disinfecting, or as solvents of adhesives, should be allowed to evaporate before the application if HF surgery. There is a risk of pooling flammable solutions under the patient or in body depressions such as the umbilicus, and in body cavities such as the vagina. Any fluids pooled in these areas should be mopped up before H.F. surgical

equipment is used. Attention should be called to the danger of ignition of endogenous gases. Some materials, for example cotton, wool and gauze, when saturated with oxygen may be ignited by sparks produced in Normal Use of the HF surgical equipment.

No modification of this equipment is allowed.

Electric Shock Hazard - Connect the power cord to a properly polarized and grounded power source with the frequency and voltage characteristics that match those listed on the back of the unit. Do not use power plug adapters.

To avoid risk of electric shock, this equipment must only be connected to a supply mains with protective earth.

Active cord removal during activation could result in a shock to the operator at the generator connector plug interface should activation occur by footswitch.

Electric Shock Hazard - Always turn off and unplug the generator before cleaning.

Fire Hazard - Do not use extension cords.

Patient Safety - Use the generator only if the self-test has been completed as described. Otherwise, inaccurate power outputs may result.

Do not use electrosurgical equipment unless properly trained to use it in the specific procedure being undertaken. Use by physicians without such training has resulted in serious, unintended patient injury, including bowel perforation and unintended, irreversible tissue necrosis.

Failure of the HF SURGICAL EQUIPMENT could result in an unintended increase of output power.

Use the lowest output setting necessary to achieve the desired surgical effect. Use the active electrode only for the minimum time necessary in order to lessen the possibility of unintended burn injury. Pediatric applications and/or procedures performed on small anatomic structures may require reduced power settings. The higher the current flow, and the longer the current is applied, the greater the possibility of unintended thermal damage to tissue, especially during use on small structures.

Avoid using power settings that would exceed the highest maximum voltage that is acceptable for each accessory. Choose only accessories that will withstand each mode and power setting.

Use of the RF Electrosurgical Generator at minimal power setting to get the expected clinical effect and for a normal clinical procedure time will not cause a surface skin temperature under the Bovie ESRS or ESRC patient return pads to rise above 41°C when the skin is prepared properly and the pad is attached properly. However be aware that extended surgical times particularly at high power will cause a continued temperature rise at the skin and return pad interface due to RF current return to the generator.

The output power selected should be as low as possible for the intended purpose. Certain devices or accessories may present a safety hazard at low power settings.

Apparent low output or failure of the A1250 RF to function correctly at the normal operating settings may indicate faulty application of the neutral electrode or poor contact in its connections. In this case, the application of the neutral electrode and its connections should be checked before selecting a higher output power.

To avoid incompatibility and unsafe operation, use suitable cables, accessories, active and neutral electrodes, including values for the highest allowed H.F. peak voltage.

Connected accessories need be rated for at least the maximum peak output voltage of the H.F. generator set at the intended output control setting in the intended operating mode.

For surgical procedures where the high frequency current could flow through parts of the body having a relatively small cross-sectional area, the use of bipolar techniques may be desirable to avoid unwanted coagulation.

For all Monopolar modes, any associated equipment and active electrodes must be rated to withstand the combination of output voltage, vp-p and crest factor as stated in Appendix A of this manual.

Associated equipment and accessories used must be rated to withstand the combination of the Vpeak rating and Crest Factor for the following RF modes, Blend, Pinpoint and Spray. When using

Cut mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 1250 V_{peak} max.

When using Blend mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 1850 V_{peak} max.

When using Coagulation mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 3300 V_{peak} max.

When using Fulguration mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 3900 V_{peak} max.

When using Bipolar mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 1200 V_{peak} max.

Use electrosurgery with caution in the presence of internal or external devices such as pacemakers or pulse generators. Interference produced by the use of electrosurgical devices can cause devices such as pacemakers to enter an asynchronous mode or can block the pacemaker effect entirely. Consult the device manufacturer or hospital Cardiology Department for further information when use of electrosurgical appliances is planned for patients with cardiac pacemakers or other implantable devices.

If the patient has an Implantable Cardioverter Defibrillator (ICD), contact the ICD manufacturer for instructions before performing an electrosurgical procedure. Electrosurgery may cause multiple activation of ICDs.

The entire area of the neutral electrode should be reliably attached to the patient's body and as close to operating field as possible. Refer to instructions for use.

The PATIENT should not come into contact with metal parts which are earthed or which have an appreciable capacitance to earth (for example operating table supports, etc.).

In some circumstances, potential exists for alternate site burns at points of skin contact (e.g., between the arm and the side of the body). This occurs when electrosurgical current seeks a path to the patient return electrode that includes the skin-to-skin contact point. Current passing through small skin-to-skin contact points is concentrated and may cause a burn. This is true for grounded, ground referenced, and isolated output generators.

When HF SURGICAL EQUIPMENT and physiological monitoring equipment are used simultaneously on the same PATIENT, any monitoring electrodes should be placed as far as possible from the surgical electrodes. In all cases, monitoring systems incorporating HIGH FREQUENCY current limiting devices are recommended.

To reduce the potential for alternate site burns, do one or more of the following:

- Avoid skin-to-skin contact points, such as fingers touching leg, when positioning the patient.
- Place 5 to 8 cm (2 to 3 in.) of dry gauze between contact points to ensure that contact does not occur.
- Position the patient return electrode to provide a direct current route between the surgical site and the return electrode which avoids skin-to-skin contact areas.
- In addition, place patient return electrodes according to the manufacturer's instructions.

Potential for alternate site burns increases if the return electrode is compromised. Bovie Medical recommends the use of split patient return electrodes and Bovie Medical generators with a contact quality monitoring system.

Do not wrap the accessory cords or patient return electrode cords around metal objects. This may induce currents that could lead to shocks, fires, or injury to the patient or surgical team.

The PATIENT leads should be positioned in such a way that contact with the PATIENT or other leads is avoided.

Minor neuromuscular stimulation is possible when arcs between the ACTIVE ELECTRODE and tissue occur. The generator has been designed to minimize the possibility of neuromuscular stimulation.

Accessories must be connected to the proper receptacle type. In particular, bipolar accessories must be connected to the Bipolar Instrument output jack only. Improper connection may result in inadvertent generator activation.

Some accessories have multiple buttons that can deliver different surgical effects. Verify accessory features and proper mode settings prior to activation.

The output power selected should be as low as possible for the intended purpose. Certain devices or ACCESSORIES may present an unacceptable RISK at low power settings.

Unless a compatible MONITORING NE is used with a CONTACT QUALITY MONITOR, loss of safe contact between the NE and the PATIENT will not result in an auditory alarm.

The generator is equipped with a return electrode sensing and contact quality monitoring system (NEM), which monitors the quality of the patient return electrode connection. When a correctly functioning single plate return electrode is connected to the generator, the NEM verifies the connections between the generator and the single return electrode. It DOES NOT verify that a single return electrode is in contact with the patient. When using a split return electrode, the NEM confirms the total resistance is within the preset safety range. Proper application (such as hydrating the patient's skin) and visual inspection of the patient return electrode is required for safe operation.

CAUTIONS

At no time should you touch the active electrode or bipolar forceps. A burn could result.

Do not stack equipment on top of the generator or place the generator on top of electrical equipment. These configurations are unstable and/or do not allow adequate cooling.

Provide as much distance as possible between the electrosurgical generator and other electronic equipment (such as monitors). An activated electrosurgical generator may cause interference with them.

Nonfunction of the generator may cause interruption of surgery. A backup generator should be available for use.

Do not turn the activation tone down to an inaudible level. The activation tone alerts the surgical team when an accessory is active.

When using a smoke evacuator in conjunction with the electrosurgical generator, place the smoke evacuator a distance from the generator and set the generator volume control at a level that ensures that the activation tones can be heard.

The use of high frequency current can interfere with the function of other electromagnetic equipment.

When high frequency surgical equipment and physiological monitoring equipment are used simultaneously on the same patient, place any monitoring electrodes as far as possible from the surgical electrodes.

Do not use needles as monitoring electrodes during electrosurgical procedures. Inadvertent electrosurgical burns may result.

To avoid the possibility of a burn to the patient, when using a split pad do not activate the unit if the solid pad indicator is illuminated green or the red alarm indicator remains illuminated red. This could indicate improper pad placement or a faulty NEM (contact quality monitor) circuit.

To avoid the possibility of an electrosurgical burn to either the patient or the physicians, do not allow the patient to come in contact with a grounded metal object during activation. When activating the unit, do not allow direct skin contact between the patient and the physician.

Remove any loose fitting jewelry from the patient before activation.

Examine all accessories and connections to the electrosurgical generator before use. Ensure that the accessories function as intended. Improper connection may result in arcs, sparks, accessory malfunction, or unintended surgical effects.

Accessories must be connected to the proper receptacle type. In particular, bipolar accessories must be connected to the Bipolar Instrument output jack only. Improper connection may result in inadvertent generator activation.

When not using active accessories, place them in a holster or in a clean, dry, nonconductive, and highly visible area not in contact with the patient. Inadvertent contact with the patient may result in burns.

Studies have shown that smoke generated during electrosurgical procedures can be potentially harmful to patients and the surgical team. These studies recommend adequately ventilating the smoke by using a surgical smoke evacuator or other means.¹

1. U.S. Department of Health and Human Services. National Institute for Occupational Safety and Health (NIOSH). Control of Smoke from Laser / Electric Surgical Procedures. HAZARD CONTROLS, Publication No. 96-128, September, 1996.

CONTRAINDICATIONS

There are no known contraindications.

NOTICES

If required by local codes, connect the generator to the hospital equalization connector with an equipotential cable.

Do not clean the generator with abrasive cleaning or disinfectant compounds, solvents, or other materials that could scratch the panels or damage the generator.

APPLICATION SPECIFICATION

Operating Conditions

RF energy is generated and passed through an interconnecting cable to an accessory where the energy is delivered to cut, coagulate and ablate tissue.

Description

- The Aaron 1250U High Frequency Electrosurgical Generators models are intended to be used for all electrosurgical cut, blend, coagulation, fulguration and bipolar procedures.

Medical Purpose / Indication

- Removal and destruction of skin lesions
- Electrosurgical cutting, blending, coagulation, fulguration and bipolar procedures of tissue to aid surgeon or physician in performing required procedures.

Site Condition

Ambient luminance range	100 lx to 1,500 lx
Viewing distance	20 cm to 200 cm
Viewing angle	normal to the display $\pm 30^\circ$

- Clean and protect from infection from start through completion of procedure.
- Note the follow Conditions of visibility for use:
 - Ambient luminance range: 100 lx to 1,500 lx
 - Viewing distance: 20 cm to 200 cm
 - Viewing angle: normal to the display $\pm 30^\circ$

Site of use

- Site of use: Tissue (ligament, cartilage)

Patient population

- Age: newborn to geriatric
- Weight: >2.5 kg
- Health: no restrictions
- Nationality: no restrictions
- Patient state: alert, relaxed maybe sedated, possible local anesthesia
 - Patient should not be User

Intended User Profile

- Education: Trained physician, physicians assistance, clinicians
 - No maximum
- Knowledge:
 - Minimum:
 - o understands electrosurgery and electrosurgical techniques
 - o read and understand supplied “User’s Guide” (accompanying document)
 - o understands hygiene
 - No maximum
- Language understanding:
 - Languages as specified in the marketing distribution plan
- Experience:
 - Minimum:
 - o Some training on techniques or training under surveillance/supervision
 - o Other: no special experience needed
 - o No maximum
- Permissible impairments:
 - Mild reading vision impairment or corrected vision to 20/20
 - impaired by 40 % resulting in 60 % of normal hearing at 500 Hz to 2 kHz

CONTROLS, INDICATORS, AND RECEPTACLES

This section describes:

- The Front and Rear Panels
- Controls, Indicators, Receptacles, the Fuse Drawer, and Ports















FRONT PANEL

Figure 2 – 1 Layout of controls, indicators, and receptacles on the front panel




SYMBOLS ON THE FRONT PANEL

Refer to the following table for descriptions of symbols found on the front panel of the Aaron 1250™.

SYMBOLS	DESCRIPTION
<i>Cut Controls</i>	
	Cut Mode
	Blend Mode
<i>Coag Controls</i>	
	Coagulation Mode
	Fulguration Mode
<i>Bipolar Controls</i>	
	Bipolar Mode
<i>Indicators</i>	
	Split Return Electrode
	Solid Return Electrode
<i>Regulatory Symbology</i>	
	Mandatory: Refer to instruction manual/guide
	Defibrillator Proof Type BF Equipment
	RF Isolated – patient connections are isolated from earth at high frequency.
	Warning: Dangerous voltage.
<i>Power Switch and Handpiece Connectors</i>	
	Patient Return Electrode
	Monopolar Output
	Bipolar Output


CUT AND BLEND CONTROLS

Figure 2 – 2 Controls for the Cut and Blend Modes





COAG AND BIPOLAR CONTROLS

Figure 2 – 3 Controls for the Coagulation, Fulguration, and Bipolar Modes



INDICATORS

Figure 2 – 4 Indicators for power, return electrodes, and footswitch control



NOTICE:
Solid pad indicator only detects that a pad is connected to the unit. The unit does not monitor pad placement on the patient.


POWER SWITCH AND RECEPTACLES

Figure 2 – 5 Location of the unit power switch and front panel receptacles



Patient Return Electrode Receptacle
Accepts a standard patient return electrode plug.

Monopolar Handswitching Receptacle
Accepts standard three-pin handpieces.
Connect handswitching accessories.




Power On/Off Switch
Turns the unit on or off.

Monopolar Footswitching Receptacle
Accepts cables or adapters equipped with standard (Bovie #12) active plugs.
Connect footswitching accessories.

Bipolar Receptacle
Accepts standard cables for bipolar handpieces.

REAR PANEL

Figure 2 – 6 Layout of connectors and controls on the rear panel



SYMBOLS ON THE REAR PANEL

Refer to the following table for descriptions of symbols found on the rear panel of the Aaron 1250™.

SYMBOLS	DESCRIPTION
	Equipotential Ground Stud
	Non-ionizing Radiation
	Volume Control
	Danger - Explosion Risk If Used With Flammable Anesthetics.
	Fuse Enclosed
	* Do not dispose of this device in the unsorted municipal waste stream.
	Footswitch Input Jack
	Read Instructions Before Use
	Manufacturer
	SGS Certification Mark; Conforms to PART 1 – ANSI/AAMI ES60601-1:2005 + C1:2009 + A2:2010 + A1:2012; CAN/CSA-C22.2 No. 60601-1:08 + C2:2011 PART 2 – AAMI 60601-2-2:2009 and CAN/CSA-C22.2 No. 60601-2-2:2009

NOTICE:

* Please note that infected medical devices must be disposed of as medical/biohazard waste and cannot be included in used electronic equipment disposal/recycling programs. In addition, certain electronic products must be returned directly to Bovie Medical Corporation. Contact your Bovie® sales representative for return instructions.

Section 3

GETTING STARTED

This section includes the following information:

- Initial Inspection
- Installation
- Function Checks
- Confirming Modes
- Performance Checks

INITIAL INSPECTION

When you first unpack your Aaron 1250™, inspect it visually:

- Look for any signs of damage.
- Verify that the shipping package contains all items listed on the packing list.

If the unit or any accessories are damaged, notify Bovie Medical Corporation's Customer Service immediately. Do not use any damaged equipment.

INSTALLATION

Place the Aaron 1250™ on any flat surface with a tilt angle not more than 10°. The unit relies on natural convection cooling. Do not block its bottom or rear vents. Ensure that air flows freely on all sides of the unit.

WARNING:

Connect the power cord to a properly polarized and grounded power source with the frequency and voltage characteristics that match those listed on the back of the unit.

FUNCTION CHECKS

Upon initial installation of the unit, perform the tests listed below. Refer to the figures in the previous chapter for the location of connectors and controls.

WARNING:

At no time should you touch the active electrode or bipolar forceps. A burn could result.

Active cord removal during activation could result in a shock to the operator at the generator connector plug interface should activation occur by footswitch.

Setting Up the Unit

1. Verify that the Power Switch is in the Off position and that no accessories are connected to the unit.
2. Connect a hospital grade power cable to the AC power cable receptacle on the back of the unit, then to a properly grounded wall outlet.
3. Connect a two-pedal footswitch to the appropriate receptacle on the back of the unit. Use only Bovie® Medical footswitches. Although other types of footswitches may fit, they may not be compatible.
4. Do not connect a patient return electrode at this time.
5. Turn the unit on by switching the power switch to the On position.

Checking the Return Electrode Alarm

1. Adjust the power settings for each mode (Cut, Blend, Coagulation, Fulguration, and Bipolar) to one watt.
2. Press the Cut pedal of the footswitch. Verify that an alarm sounds for three seconds and the Patient Return Electrode Sensing Alarm Indicator light illuminates, indicating that no return electrode is connected to the unit.
3. Verify that adjusting the volume control on the back of the unit while the alarm is sounding does not change the alarm volume.

CONFIRMING MODES

Confirm that you can select each mode and adjust the power up and down.

Checking Bipolar Mode (with footswitch)

1. Select Bipolar mode by pressing the Bipolar Mode selector.
2. Select Bipolar footcontrol by pressing the Footcontrol selector.
3. Verify that the Bipolar mode indicator illuminates and that the system generates the Coag tone when you press the Coag pedal (Blue) on the footswitch.
4. While activating the Bipolar mode, rotate the volume control over the full range to verify that the sound is audible throughout the range.
5. Confirm that releasing the Coag pedal returns the unit to an idle state.

Checking Monopolar Mode (with footswitch)

1. Select monopolar footcontrol by pressing the Footswitch Control selector until the Monopolar Footswitch Control indicator illuminates.
2. Connect a single-plate return electrode to the return electrode receptacle. Verify that the green single-plate return electrode indicator illuminates.
3. Press the Cut pedal (yellow) on the footswitch. Verify that the Cut and Blend mode activation indicator illuminates and that the system generates the Cut activation tone.
4. While activating the Cut mode, rotate the volume control over the full range to verify that the sound is audible throughout the range.
5. Press the Coag pedal (blue) on the footswitch. Verify that the Coag, Fulguration, and Bipolar activation indicator illuminates and that the system generates the Coag activation tone.
6. While activating the Coag mode, rotate the volume control over the full range to verify that the sound is audible throughout the range.

Checking Monopolar Mode (with handswitch)

1. Connect a handswitching handpiece to the Monopolar handpiece receptacle.
2. Activate, one at a time, the Cut and Coag handswitching controls. Verify that each control causes the correct indicator and tone to sound.

PERFORMANCE CHECKS

After the unit has passed the preliminary functional test, it is ready for performance testing. A qualified biomedical engineer who is thoroughly familiar with electrosurgical devices should conduct this testing. The testing should include checking all modes of operation for proper function and power output.

Section 4

USING THE AARON 1250™

This section contains the following procedures:

- Inspecting the Generator and Accessories
- Setup Safety
- Setting Up
- Preparing for Monopolar Surgery
- Preparing for Bipolar Surgery
- Activation Safety
- Activating the Unit

.....

CAUTIONS:

Read all warnings, cautions, and instructions provided with this generator before use.

Read the instructions, warnings, and cautions provided with electrosurgical accessories before use. Specific instructions are not included in this manual.

.....

INSPECTING THE GENERATOR AND ACCESSORIES

Before each use of the Aaron 1250™, verify that the unit and all accessories are in good working order:

- Inspect for damage to the Electrosurgical Generator and all its connections.
- Verify that the appropriate accessories and adapters are present.
- Inspect all cords and connectors for signs of wear, damage, and abrasion.
- Verify that no errors occur when you turn on the unit.

SETUP SAFETY

WARNINGS:

Hazardous Electrical Output - This equipment is for use only by trained, licensed physicians.

Electric Shock Hazard - Connect the generator power cord to a properly grounded receptacle. Do not use power plug adapters.

Connect the power cord to a properly polarized and grounded power source with the frequency and voltage characteristics that match those listed on the back of the unit.

No modification of this equipment is allowed.

Fire Hazard - Do not use extension cords.

Patient Safety - Use the generator only if the self-test has been completed as described. Otherwise, inaccurate power outputs may result.

The instrument receptacles on this generator are designed to accept only one instrument at a time. Do not attempt to connect more than one instrument at a time into a given receptacle. Doing so will cause simultaneous activation of the instruments.

Active cord removal during activation could result in a shock to the operator at the generator connector plug interface should activation occur by footswitch.

Use the lowest output setting necessary to achieve the desired surgical effect. Use the active electrode only for the minimum time necessary in order to lessen the possibility of unintended burn injury. Pediatric applications and/or procedures performed on small anatomic structures may require reduced power settings. The higher the current flow, and the longer the current is applied, the greater the possibility of unintended thermal damage to tissue, especially during use on small structures.

Avoid using power settings that would exceed the highest maximum voltage that is acceptable for each accessory. Choose only accessories that will withstand each mode and power setting.

Use of the RF Electrosurgical Generator at minimal power setting to get the expected clinical effect and for a normal clinical procedure time will not cause a surface skin temperature under the Bovie ESRS or ESRC patient return pads to rise above 41°C when the skin is prepared properly and the pad is attached properly. However be aware that extended surgical times particularly at high power will cause a continued temperature rise at the skin and return pad interface due to RF current return to the generator.

To avoid incompatibility and unsafe operation, use suitable cables, accessories, active and neutral electrodes, including values for the highest allowed H.F. peak voltage.

Some accessories have multiple buttons that can deliver different surgical effects. Verify accessory features and proper mode settings prior to activation.

Connected accessories need be rated for at least the maximum peak output voltage of the H.F. generator set at the intended output control setting in the intended operating mode.

Associated equipment and accessories used must be rated to withstand the combination of the Vpeak rating and Crest Factor for the following RF modes, Blend, Pinpoint and Spray.

Use electrosurgery with caution in the presence of internal or external devices such as pacemakers or pulse generators. Interference produced by the use of electrosurgical devices can cause devices such as pacemakers to enter an asynchronous mode or can block the pacemaker effect

entirely. Consult the device manufacturer or hospital Cardiology Department for further information when use of electrosurgical appliances is planned for patients with cardiac pacemakers or other implantable devices.

If the patient has an Implantable Cardioverter Defibrillator (ICD), contact the ICD manufacturer for instructions before performing an electrosurgical procedure. Electrosurgery may cause multiple activation of ICDs.

Do not use electrosurgical equipment unless properly trained to use it in the specific procedure being undertaken. Use by physicians without such training has resulted in serious, unintended patient injury, including bowel perforation and unintended, irreversible tissue necrosis.

For surgical procedures where the high frequency current could flow through parts of the body having a relatively small cross-sectional area, the use of bipolar techniques may be desirable to avoid unwanted coagulation.

In some circumstances, potential exists for alternate site burns at points of skin contact (e.g., between the arm and the side of the body). This occurs when electrosurgical current seeks a path to the patient return electrode that includes the skin-to-skin contact point. Current passing through small skin-to-skin contact points is concentrated and may cause a burn. This is true for grounded, ground referenced, and isolated output generators.

When HF SURGICAL EQUIPMENT and physiological monitoring equipment are used simultaneously on the same PATIENT, any monitoring electrodes should be placed as far as possible from the surgical electrodes. In all cases, monitoring systems incorporating HIGH FREQUENCY current limiting devices are recommended.

To reduce the potential for alternate site burns, do one or more of the following:

- Avoid skin-to-skin contact points, such as fingers touching leg, when positioning the patient.
- Place 5 to 8 cm (2 to 3 in.) of dry gauze between contact points to ensure that contact does not occur.
- Position the patient return electrode to provide a direct current route between the surgical site and the return electrode which avoids skin-to-skin contact areas.
- In addition, place patient return electrodes according to the manufacturer's instructions.

Potential for alternate site burns increases if the return electrode is compromised. Bovie Medical recommends the use of split patient return electrodes and Bovie Medical generators with a contact quality monitoring system.

Do not wrap the accessory cords or patient return electrode cords around metal objects. This may induce currents that could lead to shocks, fires, or injury to the patient or surgical team.

Minor neuromuscular stimulation is possible when arcs between the ACTIVE ELECTRODE and tissue occur. The generator has been designed to minimize the possibility of neuromuscular stimulation.

Accessories must be connected to the proper receptacle type. In particular, bipolar accessories must be connected to the Bipolar Instrument output jack only. Improper connection may result in inadvertent generator activation.

The use of flammable anesthetics or oxidizing gases such as nitrous oxide (N₂O) and oxygen should be avoided if a surgical procedure is carried out in the region of the thorax or the head, unless these agents are sucked away.

Non-flammable agents should be used for cleaning and disinfection wherever possible.

Flammable agents used for cleaning or disinfecting, or as solvents of adhesives, should be allowed to evaporate before the application if HF surgery. There is a risk of pooling flammable solutions under the patient or in body depressions such as the umbilicus, and in body cavities such as the vagina. Any fluids pooled in these areas should be mopped up before HF surgical equipment is used. Attention should be called to the danger of ignition of endogenous gases. Some materials, for example cotton, wool and gauze, when saturated with oxygen may be ignited by sparks produced in Normal Use of the HF surgical equipment.

The generator is equipped with a return electrode sensing and contact quality monitoring system (NEM), which monitors the quality of the patient return electrode connection. When a correctly functioning single plate return electrode is connected to the generator, the NEM verifies the connections between the generator and the single return electrode. It DOES NOT verify that a single

return electrode is in contact with the patient. When using a split return electrode, the NEM confirms the total resistance is within the preset safety range. Proper application and visual inspection of the patient return electrode is required for safe operation.

CAUTIONS:

Do not stack equipment on top of the generator or place the generator on top of electrical equipment. These configurations are unstable and/or do not allow adequate cooling.

Provide as much distance as possible between the electrosurgical generator and other electronic equipment (such as monitors). An activated electrosurgical generator may cause interference with them.

Non-function of the generator may cause interruption of surgery. A backup generator should be available for use.

Do not turn the activation tone down to an inaudible level. The activation tone alerts the surgical team when an accessory is active.

When using a smoke evacuator in conjunction with the electrosurgical generator, place the smoke evacuator a distance from the generator and set the generator volume control at a level that ensures that the activation tones can be heard.

NOTICES:

If required by local codes, connect the generator to the hospital equalization connector with an equipotential cable.

Connect the power cord to a wall outlet having the correct voltage. Otherwise, product damage may result.

SETTING UP

1. Verify that the generator is Off by pressing the power switch Off (O).
2. Place the generator on a stable flat surface, such as a table, platform, or medical cart. Carts with conductive wheels are recommended. For details, refer to the procedures for your institution or to local codes. Provide at least 10 to 15 cm (4 to 6 in.) of space from the sides and top of the generator for cooling. Normally, the top, sides, and rear panel are warm when you use the generator continuously for extended periods of time.
3. Plug the generator power cord into the AC Power Cable Receptacle on the rear panel.
4. Plug the generator power cord into a grounded receptacle.
5. Turn on the generator by pressing the power switch On (I). Verify the following:
 - All visual indicators and displays on the front panel illuminate.
 - Activation tones sound to verify that the speaker is working properly.
6. If the *self-test* is successful, a tone sounds. Verify the following:
 - A Cut mode is selected; a Coag or Bipolar mode is selected.
 - Each display shows a power setting. The unit automatically powers up to the Cut and Coag modes and their last selected power settings. If the Blend, Fulguration or Bipolar mode is selected, the unit will default to that modes' last set power setting.
 - The Patient Return Electrode Alarm Indicator illuminates red.

If the *self-test* is not successful, an alarm tone sounds. An error code will appear in the Cut and/or Coag display, in most cases, the generator is disabled. Note the error code and refer to Section 6, Troubleshooting.

Once the *self-test* is successful, connect the accessories and set the generator controls. Refer to Preparing for Monopolar Surgery or Preparing for Bipolar Surgery later in this section.

PREPARING FOR MONOPOLAR SURGERY

Monopolar surgery requires a return electrode.

Applying the Return Electrode

Bovie Medical recommends using return electrode contact quality monitoring system (Bovie NEM™) patient return electrodes to maximize patient safety. Using a patient return electrode without the Bovie NEM™ safety feature may result in a patient burn.

NOTICES:

The Bovie NEM™ system recommends that you use a split-plate patient return electrode.

Before activation, pad placement and visual verification of the split return electrode (split pad) indicator on the front panel is recommended. After connecting the split pad to the generator and placing the split pad securely to the patient, give the unit 5 to 10 seconds to recognize the split pad. The split pad indicator will illuminate green. If the split pad and cord are attached to the generator without secure contact to the patient, the alarm indicator will illuminate red.

Refer to the manufacturer's instructions for application site and placement procedures. When using metal plate return electrodes, use a conductive gel specifically designed for electrosurgery. Select a return electrode site with good blood flow. While a properly applied electrode results in minimal tissue heating beneath the electrode, a good blood flow helps carry heat away from the site.

Connect the cable to the Return Electrode receptacle on the front of the unit.

The unit will automatically sense the presence of a split or solid return electrode and, if a split return electrode is used, will constantly monitor the resistance at the contact between the electrode and the patient.

Connecting Accessories

1. Connect a monopolar active electrode to the unit.

If you are using...	Connect it to...
Standard 3-pin handswitching pencil	Monopolar handswitching receptacle
Footswitching pencil	Monopolar footswitching receptacle

NOTICE:

The Bovie NEM™ system recommends that you use a split return electrode.

2. If using a footswitch activated device, connect an appropriate Bovie Medical footswitch to the footswitch connecting socket on the rear of the unit.

PREPARING FOR BIPOLAR SURGERY

NOTICE:

Bipolar surgery does not require a return electrode.

1. Select the Bipolar Mode by pressing the Bipolar Mode selector. The Bipolar Mode indicator will illuminate.
2. Select Bipolar Foot Control by pressing the Foot Control selector.
3. Connect a Bipolar cable to the Bipolar receptacle.
4. Connect the appropriate Bovie Medical footswitch to the Footswitch receptacle on the rear of the unit.
5. Connect a forceps instrument to the Bipolar cable.

ACTIVATION SAFETY

WARNINGS:

Do not wrap the accessory cords or patient return electrode cords around metal objects. This may induce currents that could lead to shocks, fires, or injury to the patient or surgical team.

Minor neuromuscular stimulation is possible when arcs between the ACTIVE ELECTRODE and tissue occur. The generator has been designed to minimize the possibility of neuromuscular stimulation.

Danger: Fire / Explosion Hazard - Do not use the Aaron 1250™ in the presence of flammable anesthetics.

Fire / Explosion Hazard - The following substances will contribute to increased fire and explosion hazards in the operating room:

- Flammable substances (such as alcohol based skin prepping agents and tinctures)
- Naturally occurring flammable gases that may accumulate in body cavities such as the bowel
- Oxygen enriched atmospheres
- Oxidizing agents (such as nitrous oxide [N₂O] atmospheres).

The sparking and heating associated with electrosurgery can provide an ignition source. Observe fire precautions at all times. When using electrosurgery in the same room with any of these substances or gases, prevent their accumulation or pooling under surgical drapes, or within the area where electrosurgery is performed.

Use the lowest output setting necessary to achieve the desired surgical effect. Use the active electrode only for the minimum time necessary in order to lessen the possibility of unintended burn injury. Pediatric applications and/or procedures performed on small anatomic structures may require reduced power settings. The higher the current flow, and the longer the current is applied, the greater the possibility of unintended thermal damage to tissue, especially during use on small structures.

Use electrosurgery with caution in the presence of internal or external devices such as pacemakers or pulse generators. Interference produced by the use of electrosurgical devices can cause devices such as pacemakers to enter an asynchronous mode or can block the pacemaker effect entirely. Consult the device manufacturer or hospital Cardiology Department for further information when use of electrosurgical appliances is planned for patients with cardiac pacemakers or other implantable devices.

The output power selected should be as low as possible for the intended purpose. Certain devices or accessories may present a safety hazard at low power settings.

Apparent low output or failure of the A1250 RF to function correctly at the normal operating settings may indicate faulty application of the neutral electrode or poor contact in its connections. In this case, the application of the neutral electrode and its connections should be checked before selecting a higher output power.

When using Cut mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 1250 Vpeak max.

When using Blend mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 1850 Vpeak max.

When using Coagulation mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 3300 Vpeak max.

When using Fulguration mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 3900 Vpeak max.

When using Bipolar mode, associated equipment and active accessories should be selected that have a rated accessory voltage equal to or greater than 1200 Vpeak max.

CAUTIONS:

The use of high frequency current can interfere with the function of other electromagnetic equipment.

When high frequency surgical equipment and physiological monitoring equipment are used simultaneously on the same patient, place any monitoring electrodes as far as possible from the surgical electrodes.

Do not use needles as monitoring electrodes during electrosurgical procedures. Inadvertent electrosurgical burns may result.

To avoid the possibility of a burn to the patient, when using a split pad do not activate the unit if the solid pad indicator is illuminated green or the red alarm indicator remains illuminated red. This could indicate improper pad placement or a faulty NEM circuit.

To avoid the possibility of an electrosurgical burn to either the patient or the physicians, do not allow the patient to come in contact with a grounded metal object during activation. When activating the unit, do not allow direct skin contact between the patient and the physician.

Remove any jewelry from the patient before activation.

Studies have shown that smoke generated during electrosurgical procedures can be potentially harmful to patients and the surgical team. These studies recommend adequately ventilating the smoke by using a surgical smoke evacuator or other means.¹

Examine all accessories and connections to the electrosurgical generator before use. Ensure that the accessories function as intended. Improper connection may result in arcs, sparks, accessory malfunction, or unintended surgical effects.

Accessories must be connected to the proper receptacle type. In particular, bipolar accessories must be connected to the Bipolar Instrument output jack only. Improper connection may result in inadvertent generator activation.

When not using active accessories, place them in a holster or in a clean, dry, non-conductive, and highly visible area not in contact with the patient. Inadvertent contact with the patient may result in burns.

1. U.S. Department of Health and Human Services. National Institute for Occupational Safety and Health (NIOSH). Control of Smoke from Laser / Electric Surgical Procedures. HAZARD CONTROLS, Publication No. 96-128, September, 1996.

ACTIVATING THE UNIT

When you turn on your unit, remember this feature:

- The A1250™ automatically powers up to the Cut and Coag modes and their last selected power settings. If the Blend, Fulguration or Bipolar mode is selected, the unit will default to that modes' last set power setting.
1. Monopolar Cut - Select the mode of operation for cut: Cut or Blend, then Select the desired Cut power settings by rotating the Cut and Blend Power Control Dial.
 2. Monopolar Coag - Select the mode of operation for coagulation: Coagulation or Fulguration, then Select the coagulation power settings by rotating the Coagulation, Fulguration, and Bipolar Power Control Dial.
 3. Bipolar - Select the mode of operation for Bipolar, then select the Bipolar power settings by rotating the Coagulation, Fulguration, and Bipolar Power Control Dial.

4. Activate the generator by pressing the appropriate button:

To Activate...	Press This...	On This Device...
Monopolar		
Cut or Blend Modes	Yellow Button Yellow Pedal	Handswitching Pencil Footswitch
Coagulation or Fulguration Modes	Blue Button Blue Pedal	Handswitching Pencil Footswitch
Bipolar		
Any Bipolar	Blue (Coag) Pedal	Footswitch

NOTICES:

One footswitch activates either monopolar or bipolar footswitching accessories.

The Aaron 1250™ Electrosurgical Generator powers up to the Cut and Coag modes and their last set power settings. For example, you select the Cut mode and set the power to 50 watts. When you turn the unit off, it will default to Cut mode at 50 watts when its turned on again. The Coagulation mode will also default to the last set power setting. For the Blend, Fulguration or Bipolar modes and power settings, the unit will remember the last set power once the mode is again selected. For example, you select the Blend mode and set the power to 40 watts and select the Bipolar mode and set the power to 10 watts. When you turn the unit off, it will default to the Cut and Coag modes and their last selected power settings when its turned on again. If you then select Blend, the power will automatically default to 40 watts since it was your last set power setting.

Section 5

MAINTAINING THE AARON 1250™

This section covers the following topics:

- Cleaning
- Periodic Inspection
- Fuse Replacement

Bovie Medical Corporation recommends that you complete periodic inspection and performance testing. Perform inspections and performance testing every six months. A qualified biomedical technician should conduct this testing to ensure that the unit is operating effectively and safely.

CLEANING

After each use, clean the unit.

WARNING:

Electric Shock Hazard - Always turn off and unplug the generator before cleaning.

NOTICE:

Do not clean the generator with abrasive cleaning or disinfectant compounds, solvents, or other materials that could scratch the panels or damage the generator.

1. Turn off the generator, and unplug the power cord from the wall outlet.
2. Thoroughly wipe all surfaces of the generator and power cord with a mild cleaning solution or disinfectant and a damp cloth. Follow the procedures approved by your institution or use a validated infection control procedure. Do not allow fluids to enter the chassis. Do not sterilize the generator.

PERIODIC INSPECTION

Every six months, visually inspect the Aaron 1250™ for signs of wear or damage.

In particular, look for any of the following problems:

- Damage to the power cord
- Damage to the power cable receptacle
- Obvious damage to the unit
- Damage to any receptacle
- Accumulation of lint or debris in or around the unit.

FUSE REPLACEMENT

Fuses for the unit reside directly below the Power Cable Receptacle on the rear of the unit.

To replace the fuses, follow this procedure:

1. Unplug the power cord from the wall outlet.
2. Remove the power cord from the Power Cable Receptacle on the rear panel.
3. To release the fuse drawer, insert a small flathead screwdriver into the slot on the drawer below the power cord receptacle. Then, slide the drawer out.
4. Remove the two fuses (T3.15AL250V) and replace them with new fuses with the same values.
5. Insert the fuse holder into the Power Cable Receptacle.

NOTICE:

If the unit does not display an error and does not power on, check fuses.




Figure 5 - 1 Fuse holder

Section 6

TROUBLESHOOTING

This section includes Error Code Descriptions and actions to take to resolve them.

The Aaron 1250™ includes automatic self-diagnostics. If the diagnostics detect an error, the system displays an error code, sounds an audible tone, and deactivates the unit output power.

Most error codes result from faults in accessories attached to the unit. The following table lists the error codes, describes the errors, and recommends actions to take to resolve the errors.

If the unit displays any other error code, it requires service.

Error Code	Description	Recommended Action
F1 (on the Cut / Blend display)	Handswitch or monopolar footswitch cut pedal may be stuck.	<ol style="list-style-type: none"> 1. Turn off, then turn on the generator. Do not press buttons or activate accessory devices during the self-test. 2. If the error code reappears, disconnect all accessories. Turn off, then turn on the generator again. 3. If the problem persists, replace the handpiece or footswitch and repeat the restart. 4. If the error code reappears, record the number and call Bovie® customer service.
F1 (on the Coagulation / Fulguration Bipolar display)	Handswitch or monopolar footswitch coag pedal may be stuck.	
F2	Cut and Coag buttons activated simultaneously (pencil or footswitch)	The unit does not allow simultaneous activation of the cut and coagulation modes. Release either the cut or coag button on the handpiece, or the cut or coag pedal on the footswitch.
F3	Footswitch Cut or Coag pedal pressed while in Bipolar Foot Control and the unit is not in Bipolar Mode.	The unit will not allow the footswitch to activate the unit if Bipolar footcontrol is selected, but the Bipolar Mode is not selected.
E4	DC voltage error	<ol style="list-style-type: none"> 1. Turn the unit off. 2. Turn the unit on. 3. If the error code reappears, record the number and contact Bovie® customer service.
E6	Delta error	<ol style="list-style-type: none"> 1. Turn the unit off. 2. Verify that the unit is connected to the line voltage. 3. If the error code reappears, record the number and contact Bovie® customer service.
E7	Internal temperature of the unit exceeded the limit.	<ol style="list-style-type: none"> 1. Turn the unit off. 2. Allow the unit to cool for 20 minutes. 3. Turn the unit on. 4. If the error code reappears, record the number and contact Bovie® customer service.
E8	Connector Sense Error The unit shall monitor the connection of the main cable between the main and display boards.	<ol style="list-style-type: none"> 1. If this cable becomes disconnected an E8 error shall occur and be displayed. 2. The unit cannot be activated while the error condition is present. 3. The unit has to be reset to remove the error condition.

NOTICE:

If the unit does not power on to display an error, check fuses as described in Section 5 of this guide.

Section 7

REPAIR POLICY AND PROCEDURES

Refer to this section for information on:

- Responsibility of the Manufacturer
- Returning the Generator for Service

RESPONSIBILITY OF THE MANUFACTURER

Bovie Medical Corporation is responsible for the safety, reliability, and performance of the generator only under the following circumstances:

- The user has followed the Installation and Setup Procedures in this User's Guide.
- Persons authorized by Bovie Medical performed assembly operation, readjustments, modifications, or repairs.
- The electrical installation of the relevant room complies with local codes and regulatory requirements.
- Equipment use is in accordance with the Bovie Medical instructions for use.

For warranty information, refer to Appendix B - Warranty.

RETURNING THE GENERATOR FOR SERVICE

Before you return the generator, call your Bovie Medical representative for assistance. If instructed to send the generator to Bovie Medical Corporation, first obtain a Returned Goods Authorization Number. Then, clean the Generator and package securely to ensure proper protection of the unit. So as to aid in the processing of the unit, please be sure to include a reference to the Return Goods Authorization Number on the outside of the box and ship directly to Bovie Medical Corporation.

Step 1 – Obtain a Returned Goods Authorization Number

Call the Bovie Medical Corporation Customer Service Center to obtain a Returned Goods Authorization Number. Have the following information ready when you call:

- Hospital / clinic name / customer number
- Telephone number/fax number
- Department / address, city, state, and zip code
- Model number/Serial number
- Description of the problem
- Type of repair to be done
- P.O. number

Step 2 – Clean the Generator

WARNING:

Electric Shock Hazard - Always turn off and unplug the generator before cleaning.

NOTICE:

Do not clean the generator with abrasive cleaning or disinfectant compounds, solvents, or other materials that could scratch the panels or damage the generator.

- A. Turn off the generator, and unplug the power cord from the wall outlet.
- B. Thoroughly wipe all surfaces of the generator and power cord with a mild cleaning solution or disinfectant and a damp cloth. Follow the procedures approved by your institution or use a validated infection control procedure. Do not allow fluids to enter the chassis. You cannot sterilize the generator.

Step 3 – Ship the Generator

- A. Attach a tag to the generator that includes the Returned Goods Authorization Number and the information (hospital, phone number, etc.) listed in Step 1 – Obtain a Returned Goods Authorization Number.
- B. Be sure the generator is completely dry before you pack it for shipment. Although the preference is to have the Generator repackaged using its original packaging, Bovie® understands that this may not always be possible. If necessary, contact Customer Service for the proper packaging to ship the unit. Please be sure to include a reference of the Bovie Return Goods Authorization Number on the outside of the box/container.
- C. Ship the generator, prepaid, to the address given to you by the Bovie Medical Corporation Service Center.

Appendix

A

TECHNICAL SPECIFICATIONS

All specifications are nominal and subject to change without notice. A specification referred to as “typical” is within $\pm 20\%$ of a stated value at room temperature (25° C / 77° F) and a nominal input power voltage.

PERFORMANCE CHARACTERISTICS

Input Power

100 – 240 VAC
Mains line frequency range (nominal): 50 – 60 Hz
Power consumption: 3A~
Fuses (two): 3.15A (Slow Blow)

Duty Cycle

Under maximum power settings and rated load conditions (Pure Cut, 120 watt @ 500 ohm load), the generator is suitable for activation times of 10 seconds on, 30 seconds off for one hour.

The internal temperature of the unit is continuously monitored. If the temperature rises above 85° C, the alarm will sound and output power will be deactivated.

Dimensions and Weight

Width	26 cm (10.25 in.)	Depth	30.5 cm (12 in.)
Height	15.2 cm (6 in.)	Weight	< 4 kg (< 9 lbs)

Operating Parameters

Ambient temperature range	10° to 40° C (50° to 104° F)
Relative humidity	30% to 75%, non-condensing
Atmospheric pressure	70kPa to 106kPa
Warm-up time	If transported or stored at temperatures outside the operating temperature range, allow one hour for the generator to reach room temperature before use.

Transport

Ambient temperature range	-40° to +70° C
Relative humidity	10% to 100%, including condensation
Atmospheric pressure	50kPa to 106kPa

Storage

Ambient temperature range	10° to 30° C (68° to 86° F)
Relative humidity	10% to 75%, non-condensing
Atmospheric pressure	50kPa to 106kPa

Audio Volume

The audio levels stated below are for activation tones (bipolar, cut and coag) and alarm tones (return electrode and system alarms) at a distance of one meter. Alarm tones meet the requirements for IEC 60601-2-2.

Activation Tone

Volume (adjustable)	40 to ≥ 65 dB
Frequency	Cut: 610 Hz \pm 10 Hz Blend: 610 Hz \pm 10 Hz Pinpoint: 840 Hz \pm 10 Hz Spray: 840 Hz \pm 10 Hz Bipolar: 840 Hz \pm 10 Hz
Duration	Continuous while the generator is activated

Alarm Tone

Volume (not adjustable)	≥ 65 dB at a distance of one meter
Frequency	2.44 kHz / 450 ms / 1.22 kHz / 450 ms

Return Electrode Sensing

The system presents audible and visible alarms when it senses no return electrode.

Single Plate	Trip resistance: 0 Ω to 8 Ω \pm 1 Ω Continuous measurement: Once the system establishes the single-plate electrode resistance, an increase of 20 Ω \pm 25 Ω in resistance will cause an alarm. When the alarm condition exists, the system deactivates output power.
Split Plate	Trip resistance: 10 Ω \pm 5 Ω to 135 Ω \pm 10 Ω Continuous measurement: Once the system establishes the split-plate electrode resistance, an increase of (35 \pm 5)% in resistance will cause an alarm. When the alarm condition exists, the system deactivates output power.

Low Frequency (50-60 Hz) Leakage Current

Enclosure source current, ground open	< 500 μ A 220 - 240 VAC
	< 300 μ A 90 - 120 VAC
Source current, patient leads, all outputs	Normal polarity, intact ground: < 10 μ A Normal polarity, ground open: < 50 μ A Reverse polarity, ground open: < 50 μ A
Sink current at high line, all inputs	< 50 μ A

High Frequency (RF) Leakage Current

Bipolar RF leakage current	< 39 mA _{rms}
Monopolar RF leakage current (additional tolerance)	< 150 mA _{rms}

Operating Conditions

RF energy is generated and passed through an interconnecting cable to an accessory where the energy is delivered to cut, coagulate and ablate tissue.

STANDARDS AND IEC CLASSIFICATIONS

Class I Equipment (IEC 60601-1)

Accessible conductive parts cannot become live in the event of a basic insulation failure because of the way in which they are connected to the protective earth conductor.



Type BF Equipment (IEC 60601-1) / Defibrillator Proof

The Aaron 1250™ Electrosurgical Generator provides a high degree of protection against electric shock, particularly regarding allowable leakage currents. It is type BF equipment. Patient connections are isolated from earth and resist the effects of defibrillator discharge.

Drip Proof (IEC 60601-2-2)

The generator enclosure is constructed so that liquid spillage in normal use does not wet electrical insulation or other components which, when wet, are likely to affect adversely the safety of the generator.

Electromagnetic Interference

When other equipment is placed on or beneath an activated Aaron 1250™ Electrosurgical Generator, the unit can be activated without interference. The generator minimizes electromagnetic interference to video equipment used in the operating room.

Electromagnetic Compatibility (IEC 60601-1-2 and IEC 60601-2-2)

The Aaron 1250™ Electrosurgical Generator complies with the appropriate IEC 60601-1-2 and IEC 60601-2-2 specifications regarding electromagnetic compatibility.

Voltage Transients (Emergency Generator Mains Transfer)

The Aaron 1250™ Electrosurgical Generator operates in a safe manner when the transfer is made between line AC and an emergency generator voltage source.

EMC COMPLIANCE

Special precautions should be taken regarding the 1250™. Medical Electrical Equipment needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided in this manual.

Understand that only the Accessories supplied with or ordered from Bovie Medical Corporation should be used with your device. The use of Accessories, transducers, and cables other than those specified, may result in increased Emissions or decreased Immunity of the A1250™. The A1250™ and its accessories are not suitable for interconnection with other equipment.

Portable and mobile RF communications equipment can affect Medical Electrical Equipment. The 1250™ should not be used adjacent to or stacked with other equipment and that if adjacent or stacked use is necessary, the A1250™ should be observed to verify normal operation in the configuration in which it will be used.

Recommended separation distances between portable and mobile RF communications equipment and the A1250™.

The 1250™ is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the A1250™ can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the A1250™ as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter W	separation distance according to frequency of transmitter in metres (m)		
	150 kHz to 80 MHz $d = \left[\frac{3.5}{V_f} \right] \sqrt{P}$	80 MHz to 800 MHz $d = \left[\frac{3.5}{E_f} \right] \sqrt{P}$	800 MHz to 2.5 GHz $d = \left[\frac{7}{E_f} \right] \sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.


NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

Guidance and manufacturer's declaration – electromagnetic emissions

The A1250™ is intended for use in the electromagnetic environment listed below. The customer or the user of the A1250™ should assure that it is used in such an environment.

Emissions test	Compliance	Electromagnetic environment - guidance
RF Emissions CISPR 11	Group 2	The A1250™ must emit electromagnetic energy in order to perform its intended function. Nearby electronic equipment may be affected.
RF Emissions CISPR 11	Class A	The A1250™ is suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used in domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations/flicker emissions IEC 61000-3-3	Complies	

Guidance and manufacturer's declaration – electromagnetic immunity			
The A1250™ is intended for use in the electromagnetic environment listed below. The customer or the user of the A1250™ should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Electrostatic discharge (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	±2 kV for power supply lines ±1 kV for input/output lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV differential mode ±2 kV common mode	±1 kV differential mode ±2 kV common	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5 % U_t (>95 % dip in U_t) for 0.5 cycle 40 % U_t (60 % dip in U_t) for 5 cycles 70 % U_t (30 % dip in U_t) for 25 cycles <5 % U_t (>95 % dip in U_t) for 5 sec	<5 % U_t (>95 % dip in U_t) for 0.5 cycle 40 % U_t (60 % dip in U_t) for 5 cycles 70 % U_t (30 % dip in U_t) for 25 cycles <5 % U_t (>95 % dip in U_t) for 5 sec	Mains power quality should be that of a typical commercial or hospital environment. If the user of the A1250™ requires continued operation during power mains interruptions, it is recommended that the A1250™ be powered from an uninterruptible power supply or a battery.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
NOTE U_t is the a.c. mains voltage prior to application of the test level.			

Guidance and manufacturer's declaration – electromagnetic immunity continued...			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms (V_1)	<p>Portable and mobile RF communications equipment should be used no closer to any part of the A1250™, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.</p> <p>Recommended separation distance</p> $d = \left[\frac{3.5}{3} \right] \sqrt{P}$
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m (E_1)	$d = \left[\frac{3.5}{3} \right] \sqrt{P}$ <p>80 MHz to 800 MHz</p> $d = \left[\frac{7}{3} \right] \sqrt{P}$ <p>800 MHz to 2.5 GHz</p> <p>where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m)</p> <p>Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey,^a should be less than the compliance level in each frequency range.^b</p> <p>Interference may occur in the vicinity of equipment marked with the following symbol.</p> 
<p>NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.</p> <p>NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.</p>			
<p>^a Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicated theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location which the A1250™ is used exceeds the applicable RF compliance level above, the A1250™ should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the A1250™.</p> <p>^b Over the frequency range 150 kHz to 80 MHz, field strengths should be less than $[V_3]$ V/m.</p>			

OUTPUT CHARACTERISTICS

Maximum Output for Monopolar and Bipolar Modes

Power readouts agree with actual power into rated load to within 20% or 5 watts, whichever is greater.

Mode	Output Power	Output Frequency	Repetition Rate	Open Circuit V_{peak max}	Crest Factor* (Rated Load)
Cut	120 W @ 500 Ω	357 kHz ± 50 kHz	N / A	1250V	2.9 ± 20%
Blend	90 W @ 800 Ω	357 kHz ± 50 kHz	30 kHz ± 5 kHz	1850V	3.3 ± 20%
Coagulation	80 W @ 1000 Ω	475 kHz ± 19 kHz	57 kHz ± 5 kHz	3300V	5.5 ± 20%
Fulguration	40 W @ 1000 Ω	410 kHz ± 50 kHz	25 kHz ± 5 kHz	3900V	7.7 ± 20%
Bipolar	30 W @ 200 Ω	520 kHz (-14 kHz, +29 kHz)	32 kHz ± 5 kHz	1200V	6.9 ± 20%

* an indication of a waveform's ability to coagulate bleeders without a cutting effect

OUTPUT POWER CURVES

Figure A-1 and A-2 illustrates output voltage (V_{peak}) versus power setting. Figure A-3 illustrates output power versus power setting for all modes. Figures A-4 through A-8 illustrate specific output power delivered to a range of load resistances for each mode.

Figure A – 1 Output voltage (V_{peak}) versus power setting (Cut, Coag)




Figure A – 2 Output voltage (V_{peak}) versus power setting (Bipolar)




Figure A – 3 Output power versus power setting for all modes




Figure A – 4 Output power vs impedance for Cut mode




Figure A – 5 Output power vs impedance for Blend mode




Figure A – 6 Output power versus impedance for Coagulation modes




Figure A – 7 Output power versus impedance for Fulguration mode





Figure A – 8 Output power vs impedance for Bipolar mode



WARRANTY

Bovie Medical Corporation, warrants each product manufactured by it to be free from defects in material and workmanship under normal use and service for the period(s) set forth below.

Bovie Medical Corporation's obligation under this warranty is limited to the repair or replacement, at its sole option, of any product, or part thereof, which has been returned to it or its Distributor within the applicable time period shown below after delivery of the product to the original purchaser, and which examination discloses, to Bovie Medical Corporation's satisfaction, that the product is indeed, defective.

This warranty does not apply to any product, or part thereof, which has been repaired or altered outside Bovie Medical Corporation's factory in a way so as, in Bovie Medical Corporation's judgment, to affect its stability or reliability, or which has been subjected to misuse, neglect, or accident.

The warranty periods for Bovie Medical products are as follows:

- Electrosurgical Generators: Four years from date of shipment
- Mounting Fixtures (all models): Two years from date of shipment
- Footswitches (all models): One year from date of shipment
- Patient Return Electrodes: Shelf life only as stated on packaging
- Sterile Single Use Accessories: Only as stated on packaging

This warranty is in lieu of all other warranties, express or implied, including without limitation, the warranties of merchantability and fitness for a particular purpose, and of all other obligations or liabilities on the part of Bovie Medical Corporation.

Bovie Medical Corporation neither assumes nor authorizes any other person to assume for it any other liability in connection with the sale or use of any of Bovie Medical's products.

Notwithstanding any other provision herein or in any other document or communication, Bovie Medical Corporation's liability with respect to this agreement and products sold hereunder shall be limited to the aggregate purchase price for the goods sold by Bovie Medical Corporation to the customer.

Bovie Medical Corporation disclaims any liability hereunder or elsewhere in connection with the sale of this product, for indirect or consequential damages.

This warranty and the rights and obligations hereunder shall be construed under and governed by the laws of the State of Florida, USA.

The sole forum for resolving disputes arising under or relating in any way to this warranty is the District Court of the County of Pinellas, State of Florida, USA.

Bovie Medical Corporation, its dealers, and representatives reserve the right to make changes in equipment built and/or sold by them at any time without incurring any obligation to make the same or similar changes on equipment previously built and/or sold by them.

MC-55-128-001 REV. 8

2016-01-19



BOVIE MEDICAL CORPORATION

5115 Ulmerton Road

Clearwater, Florida 33760 USA

U.S. Phone 1-800-537-2790 • Fax 1-800-323-1640

International Phone +1-727-384-2323 • www.boviemed.com



Emergo Europe

Molenstraat 15

2513 BH, The Hague

The Netherlands