SDPharmLabs

PharmLabs San Diego Certificate of Analysis

3421 Hancock St, Second Floor, San Diego, CA 92110 | License: C8-0000098-LIC ISO/IEC 17025:2017 Certification L17-427-1 | Accreditation #85368

Sample KO Blend Pre-Roll - Sour Diesel



Laboratory note: The estimated concentration of the unknown peak in the sample is 0.00% | Currently PharmLabs laboratory can not confirm an unidentified peak in your chromatogram due to interference (only with highly concentrated D8 products) from which we believe to be either (+)d8-THC or 49-THC. At this time there are no reference standards available for (+)d8-THC is o different compound from the main (-)d8-THC cannobinoid and, therefore, these two compounds may have different efficacies. Using the most advanced instruments and techniques available, the separation of (+)d8-THC and d9-THC is problematic for the scientific community as a whole. PharmLabs believes the unidentified peak to be a combination of (+)d8-THC with the majority, if not all, of the concentration being (+)d8-THC. Total (+/-) D8 Concentration is estimated to be 18.37%

*CAN+ - Cannabinoids Analysis

Analyzed Oct 17, 2022 | Instrument HPLC-VWD | Method SOP-001

	LOD	LOO	Result	Result
Analyte	mg/g	mg/g	%	mg/g
Cannabidivarin (CBDV)	0.039	0.16	ND	ND
Cannabidiolic Acid (CBDA)	0.001	0.16	11.92	119.16
Cannabigerol Acid (CBGA)	0.001	0.16	0.30	3.01
Cannabigerol (CBG)	0.001	0.16	0.06	0.62
Cannabidiol (CBD)	0.001	0.16	1.63	16.27
Tetrahydrocannabivarin (THCV)	0.001	0.16	<loq< td=""><td><loq< td=""></loq<></td></loq<>	<loq< td=""></loq<>
Cannabinol (CBN)	0.001	0.16	<loq< td=""><td><loq< td=""></loq<></td></loq<>	<loq< td=""></loq<>
Tetrahydrocannabinol (Δ9-THC)	0.003	0.16	UI	UI
Δ8-tetrahydrocannabinol (Δ8-THC)	0.004	0.16	18.37	183.70
Cannabicyclol (CBL)	0.002	0.16	ND	ND
Cannabichromene (CBC)	0.002	0.16	0.15	1.46
Tetrahydrocannabinolic Acid (THCA)	0.001	0.16	0.34	3.36
Total THC (THCa * 0.877 + THC)			0.29	2.94
Total CBD (CBDa * 0.877 + CBD)			12.08	120.77
Total CBG (CBGa * 0.877 + CBG)			0.33	3.26
TOTAL CANNABINOIDS			31.23	312.25

*Dru Weiaht %

MWA - Moisture Content & Water Activity Analysis

Analyzed Oct 14, 2022 | Instrument Chilled-mirror Dewpoint and Capacitance | Method SOP-008

Analyte	Result	Limit	Analyte	Result	Limit
Moisture (Moi)	7.1 % Mw	13 % Mw	Water Activity (WA)	0.51 a _w	0.85 a _w

UI Not Identified
ND Not Detected
N/A Not Applicable
NT Not Reported
LOD Limit of Detection
LOQ Limit of Quantification
<LOQ Detection
LOQ Limit of Guantification
<LOQ Detection
Forum of Countification
CEU/Q Colony Forming Units per 1 gram
TNTC Too Numerous to Count







Authorized Signature

Brandon Starr



